

# Validation of the UEBE Visomat Double Comfort upper arm blood pressure monitor, in auscultation mode, for clinic use and self measurement in a general population, according to the European Society of Hypertension International Protocol revision 2010

Serena Masiero, Claudio Fania, Paolo Palatini

**Keywords:** validation, blood pressure, device, automatic, measurement

*Department of Clinical and Experimental Medicine, University of Padova, Italy*

Correspondence to Paolo Palatini, Clinica Medica 4, via Giustiniani, 2, 35128 Padova, Italy

Tel: 0039-049-8212278; e-mail: palatini@unipd.it

## Abstract

The UEBE Visomat Double Comfort, an upper arm blood pressure monitor, in auscultation mode, for clinical use and self measurement, was validated, in a general population, according to the European Society of Hypertension International Protocol revision 2010. The protocol requirements were followed precisely. The device passed all of the requirements and, fulfilling the standards of the protocol, is recommended for clinical use.

## Device Details

Brand	UEBE
Model	Visomat Double Comfort
Manufacturer	UEBE
Location	Upper Arm
Method	Auscultation
Purpose	Clinic Measurement
Operation	Fully Automatic
Arm Cuffs	Other cuffs: 23.0 cm to 43.0 cm
Other Features	Combines auscultatory and oscillometric determination of blood pressure. The Auscultatory measurement is the reference procedure.



## Methodology

### **Familiarisation**

Forty-one test-measurements were carried out. No problems were encountered.

### **Recruitment**

Hypertensive subjects were recruited from those attending the hypertension clinic in the Medical Policlinic, University of Padua, Italy. All 42 participated immediately and without appointment. The 11 normotensive subjects were recruited from inpatients (four) and from hospital staff (seven). There was some difficulty in recruiting subjects with SBP and DBP in the high range but, other than that, there were no problems.

**Screening and Recruitment Details**

Screening and Recruitment			Recruitment Ranges			
Total Screened	42		mmHg	All	On Rx	
Total Excluded	9		< 90	1	1	
Ranges Complete	1	Low	90 - 129	10		
Ranges Adjustment	7	SBP	Medium 130 - 160	12	8	
Arrhythmias	0		High 161 - 180	9		
Device Failure	0		> 180	1	7	
Poor Quality Sounds	1					
Cuff Size Unavailable	0	Low	< 40	0	1	
Observer Disagreement	0		40 - 79	12		
Distribution	0	DBP	Medium 80 - 100	10	6	
Other Reasons	0		High 101 - 130	11		
Total Recruited	33		> 130	0	9	

**Procedure**

The European Society of Hypertension International Protocol revision 2010 for the validation of blood pressure measuring devices in adults was followed precisely.[1] Overseen by an independent supervisor, measurements were recorded by two observers blinded from both each other's readings and from the device readings.

**Results****Subject Details**

Sex			
Male : Female	18 : 15		
Age (years)			
Range (Low : High)	25 : 88		
Mean (SD)	55.0 (20.2)		
Arm Circumference (cm)			
Range (Low : High)	23.0 : 38.0		
Mean (SD)	29.1 (4.1)		
Cuff for test device			
Other	33	(23.0 - 43.0 cm)	
	SBP	DBP	
Recruitment BP (mmHg)			
Range (Low : High)	84 : 190	50 : 120	
Mean (SD)	140.0 (28.0)	86.6 (17.2)	

**Observer Measurements in each Recruitment Range**

SBP (mmHg)	DBP (mmHg)
Overall Range (Low : High)	Overall Range (Low : High)
79 : 191	49 : 122
Low (< 130)	Low (< 80)
33	41
Medium (130 – 160)	Medium (80 – 100)
42	34
High (> 160)	High (> 100)
24	24
Maximum Difference	Maximum Difference
18	17


**Observer Differences**

	SBP (mmHg)	DBP (mmHg)	Repeated measurements
Observer 2 – Observer 1			
Range (Low : High)	-2 : +4	-4 : +4	
Mean (SD)	0.2 (1.9)	0.0 (1.9)	0

**Validation Results**

Part 1	≤ 5 mmHg	≤ 10 mmHg	≤ 15 mmHg	Grade 1	Mean (mmHg)	SD (mmHg)
Pass Requirements						
Two of	73	87	96			
All of	65	81	93			
Achieved						
SBP	83	98	99	Pass	-1.1	4.0
DBP	85	97	99	Pass	-0.4	4.1
Part 2						
	2/3 ≤ 5 mmHg	0/3 ≤ 10 mmHg		Grade 2		Grade 3
Pass Requirements						
	≥ 24	≤ 3				
Achieved						
SBP	31	0		Pass		Pass
DBP	30	0		Pass		Pass
Part 3						
						Result
						<b>PASS</b>

**Plots**



**Discussion**

Recruitment of subjects with high DBP proved to be difficult and accounted for most of the extra screened subjects; this is reflected in the overall distribution, as shown in the DBP plot, in which most of the points are below 115 mmHg. On the other hand, the agreement between observer and device was similar in the three BP ranges and all BP discrepancies were within 15 mmHg.

The visomat double comfort monitor is provided with two operational modalities, a microphonic mode and an oscillometric mode. In the standard operational modality, tested in the present study, the device uses the microphonic method and when a failing auscultatory measurement occurs the oscillometric measurement is taken into account. The UEBE visomat double comfort device used in the microphonic mode fulfilled the validation criteria of the International protocol.

**Conclusion**

As the device has reached the required standards, it is recommended for clinical and personal use in a general population.

**Acknowledgements and Conflict of Interest**

Three monitors were supplied for the purposes of the study by the manufacturer who also funded the study. None of the authors has any association with UEBE or has received any personal benefit from UEBE.

## References

1. O'Brien E, Atkins N, Stergiou G, Karpettas N, Parati G, Asmar R, Imai Y, Wang J, Mengden T, Shennan A; on behalf of the Working Group on Blood Pressure Monitoring of the European Society of Hypertension. European Society of Hypertension International Protocol revision 2010 for the Validation of Blood Pressure Measuring Devices In Adults. *Blood Press Monit* 2010;15:23–38.