

# Validation of the OMRON HBP-1320 upper arm blood pressure monitor, in oscillometry mode, for clinic use in a general population, according to the European Society of Hypertension International Protocol revision 2010

Hakuo Takahashi

**Keywords:** Blood pressure, European Society of Hypertension, guideline, device, measurement

Department of Cardiology, Biwako Central Hospital, 22-33, Gotenbama, Otsu, Shiga, 520-0834, Japan

Correspondence to Hakuo Takahashi, MD, Director, Department of Cardiology, Biwako Central Hospital, 22-33, Gotenbama, Otsu, Shiga, 520-0834, Japan

Tel: +81-77-526-2131; e-mail: takahashi@kou-sei-kai.or.jp

## Abstract

The OMRON HBP-1320, an upper arm blood pressure monitor, in oscillometry mode, for clinical use, was validated, in a general population, according to the European Society of Hypertension International Protocol revision 2010. The protocol requirements were followed precisely. The device passed all of the requirements and, fulfilling the standards of the protocol, is recommended for clinical use.

## Device Details

Brand	OMRON
Model	HBP-1320
Manufacturer	OMRON HEALTHCARE Co., Ltd.
Location	Upper Arm
Method	Oscillometry
Purpose	Clinic Measurement
Operation	Fully Automatic
Arm Cuffs	Small Adult: 17.0 cm to 22.0 cm, Standard Adult: 22.0 cm to 32.0 cm, Large Adult: 32.0 cm to 42.0 cm and other cuffs: 12.0 cm to 50.0 cm
Other Features	The cuff size "Other" refers to either the Extra Small Cuff, 12.0cm to 18.0cm, or to the Extra Large cuff, 42.0cm to 50.0cm. There are functions to detect irregular pulse wave and body movement. The function to enable auscultation by an observer.



## Methodology

### Familiarisation

Numerous test-measurements were carried out. No problem was found.

### Recruitment

Hypertensive subjects were recruited from outpatients clinic in the Department of Cardiology in Biwako Central Hospital (Shiga, Japan). Some participated immediately without appointment. Normotensive subjects were recruited from outpatients and volunteers.

**Screening and Recruitment Details**

Screening and Recruitment			Recruitment Ranges		
Total Screened	64		mmHg	All	On Rx
Total Excluded	31		< 90	1	2
Ranges Complete	0		Low	90 - 129	10
Ranges Adjustment	0	SBP	Medium	130 - 160	11
Arrhythmias	0		High	161 - 180	10
Device Failure	0		> 180	1	0
Poor Quality Sounds	0				
Cuff Size Unavailable	0		Low	< 40	0
Observer Disagreement	0		40 - 79	11	2
Distribution	0	DBP	Medium	80 - 100	12
Other Reasons	31		High	101 - 130	10
Total Recruited	33		> 130	0	0

**Procedure**

The European Society of Hypertension International Protocol revision 2010 for the validation of blood pressure measuring devices in adults was followed precisely.[1] Overseen by an independent supervisor, measurements were recorded by two observers blinded from both each other's readings and from the device readings.

**Results****Subject Details**

Sex		
Male : Female	10 : 23	
Age (years)		
Range (Low : High)	31 : 73	
Mean (SD)	52.9 (9.5)	
Arm Circumference (cm)		
Range (Low : High)	16.6 : 42.1	
Mean (SD)	28.6 (5.9)	
Cuff for test device		
Small	4	(17.0 - 22.0 cm)
Standard	19	(22.0 - 32.0 cm)
Large	8	(32.0 - 42.0 cm)
Other	2	(12.0 - 50.0 cm)
	SBP	DBP
Recruitment BP (mmHg)		
Range (Low : High)	84 : 210	46 : 129
Mean (SD)	142.3 (31.3)	88.2 (20.0)

**Observer Measurements in each Recruitment Range**

SBP (mmHg)		DBP (mmHg)	
Overall Range (Low : High)	92 : 196	Overall Range (Low : High)	47 : 128
Low (< 130)	33	Low (< 80)	32
Medium (130 – 160)	30	Medium (80 – 100)	28
High (> 160)	36	High (> 100)	39
Maximum Difference	6	Maximum Difference	11


**Observer Differences**

	SBP (mmHg)	DBP (mmHg)	Repeated measurements
Observer 2 – Observer 1			
Range (Low : High)	-4 : +4	-4 : +4	
Mean (SD)	0.1 (1.5)	-0.1 (1.7)	0

**Validation Results**

Part 1	$\leq 5$ mmHg	$\leq 10$ mmHg	$\leq 15$ mmHg	Grade 1	Mean (mmHg)	SD (mmHg)
Pass Requirements						
Two of	73	87	96			
All of	65	81	93			
Achieved						
SBP	75	95	98	Pass	-0.4	4.9
DBP	83	97	99	Pass	-0.2	4.2
Part 2	$2/3 \leq 5$ mmHg	$0/3 \leq 5$ mmHg	Grade 2	Grade 3		
Pass Requirements						
	$\geq 24$	$\leq 3$				
Achieved						
SBP	28	1	Pass			Pass
DBP	30	0	Pass			Pass
Part 3						Result
						<b>PASS</b>

**Plots**



**Discussion**

The study finished without any problems. However, it was hard to recruit patients with high blood pressure levels of 161 to 180mmHg. The agreement between observer and device was similar in the three BP ranges and the magnitude of discrepancies were within 15mmHg.

**Conclusion**

As the device has reached the required standards, it is recommended for clinical and personal use in a general population.

**Acknowledgements and Conflict of Interest**

The monitor was supplied for the purposes of the study by the manufacturer OMRON Healthcare CO.,LTD. who also funded the study. The author does not have any association with OMRON Healthcare CO.,LTD. and did not receive any personal benefit from OMRON Healthcare CO.,LTD..

## References

1. O'Brien E, Atkins N, Stergiou G, Karpettas N, Parati G, Asmar R, Imai Y, Wang J, Mengden T, Shennan A; on behalf of the Working Group on Blood Pressure Monitoring of the European Society of Hypertension. European Society of Hypertension International Protocol revision 2010 for the Validation of Blood Pressure Measuring Devices In Adults. *Blood Press Monit* 2010;15:23–38.