

# Validation of the OMRON HBP-1300 upper arm blood pressure monitor, in oscillometry mode, for clinic use in a general population, according to the European Society of Hypertension International Protocol revision 2010

Hakuo Takahashi, Toyohiko Yokoi and Masamichi Yoshika

**Keywords:** blood pressure, European Society of Hypertension, guideline, measurement, device

*Department of Clinical Sciences and Laboratory Medicine, Kansai Medical University, Hirakata Osaka, Japan*

Correspondence to Professor Hakuo Takahashi, MD, PhD, Department of Clinical Sciences and Laboratory Medicine, Kansai Medical University, Shin-machi 2-5-1, Hirakata Osaka 573-1010, Japan

Tel: +81 72 804 2691; e-mail: takahash@hirakata.kmu.ac.jp

## Abstract

The OMRON HBP-1300, an upper arm blood pressure monitor, in oscillometry mode, for clinical use, was validated, in a general population, according to the European Society of Hypertension International Protocol revision 2010. The protocol requirements were followed precisely. The device passed all of the requirements and, fulfilling the standards of the protocol, is recommended for clinical use.

## Device Details

Brand	OMRON
Model	HBP-1300
Manufacturer	OMRON HEALTHCARE Co., Ltd.
Location	Upper Arm
Method	Oscillometry
Purpose	Clinic Measurement
Operation	Fully Automatic
Arm Cuffs	Small Adult: 17.0 cm to 22.0 cm, Standard Adult: 22.0 cm to 32.0 cm, Large Adult: 32.0 cm to 42.0 cm and other cuffs: 12.0 cm to 50.0 cm
Other Features	The cuff size "Other" refers to either the Extra Small cuff, 12 cm to 18 cm, or to the Extra Large cuff 42 cm to 50 cm. There is a function to detect irregular pulse wave and body movement. The function to enable auscultation by an observer.



## Methodology

### Familiarisation

Hundreds of test-measurements were carried out. No problems were encountered.

### Recruitment

Hypertensive subjects were recruited from outpatients clinic in the Department of Cardiology in Kansai Medical University, Hirakata Hospital (Osaka, Japan). Some participated immediately without appointment. Normotensive subjects were recruited from outpatients and volunteers.

**Screening and Recruitment Details**

Screening and Recruitment			Recruitment Ranges		
Total Screened	51		mmHg	All	On Rx
Total Excluded	18		< 90	0	
Ranges Complete	0		Low 90 - 129	10	4
Ranges Adjustment	0	SBP	Medium 130 - 160	11	9
Arrhythmias	3		High 161 - 180	10	4
Device Failure	0		> 180	2	
Poor Quality Sounds	3				
Cuff Size Unavailable	0		< 40	0	6
Observer Disagreement	0		Low 40 - 79	11	
Distribution	0	DBP	Medium 80 - 100	12	9
Other Reasons	12		High 101 - 130	10	2
Total Recruited	33		> 130	0	

**Procedure**

The European Society of Hypertension International Protocol revision 2010 for the validation of blood pressure measuring devices in adults was followed precisely.[1] Overseen by an independent supervisor, measurements were recorded by two observers blinded from both each other's readings and from the device readings.

**Results****Subject Details**

Sex		
Male : Female	12 : 21	
Age (years)		
Range (Low : High)	34 : 88	
Mean (SD)	56.9 (14.0)	
Arm Circumference (cm)		
Range (Low : High)	15.3 : 43.0	
Mean (SD)	28.3 (5.4)	
Cuff for test device		
Small	3	(17.0 - 22.0 cm)
Standard	24	(22.0 - 32.0 cm)
Large	4	(32.0 - 42.0 cm)
Other	2	(12.0 - 50.0 cm)
	SBP	DBP
Recruitment BP (mmHg)		
Range (Low : High)	90 : 192	49 : 127
Mean (SD)	144.9 (27.2)	86.8 (19.8)

**Observer Measurements in each Recruitment Range**

SBP (mmHg)		DBP (mmHg)	
Overall Range (Low : High)	80 : 180	Overall Range (Low : High)	48 : 124
Low (< 130)	38	Low (< 80)	36
Medium (130 – 160)	38	Medium (80 – 100)	38
High (> 160)	23	High (> 100)	25
Maximum Difference	15	Maximum Difference	13


**Observer Differences**

	SBP (mmHg)	DBP (mmHg)	Repeated measurements
Observer 2 – Observer 1			
Range (Low : High)	-2 : +4	-4 : +4	
Mean (SD)	0.3 (1.3)	0.5 (1.4)	0

**Validation Results**

Part 1	$\leq 5$ mmHg	$\leq 10$ mmHg	$\leq 15$ mmHg	Grade 1	Mean (mmHg)	SD (mmHg)
Pass Requirements						
Two of	73	87	96			
All of	65	81	93			
Achieved						
SBP	81	97	99	Pass	-0.7	4.0
DBP	82	96	99	Pass	0.7	4.2
Part 2	$2/3 \leq 5$ mmHg	$0/3 \leq 10$ mmHg	Grade 2	Grade 3		
Pass Requirements						
	$\geq 24$	$\leq 3$				
Achieved						
SBP	30	0	Pass			Pass
DBP	27	0	Pass			Pass
Part 3						Result
						<b>PASS</b>

**Plots**



**Discussion**

No specific problem were encountered during validation and distribution conditions were fulfilled. But it was difficult to fulfill requirements of the International Protocol because some subjects were very hypertensive over SBP 200mmHg. The agreement between observer and device was similar in the three BP ranges and all BP discrepancies were within 15 mmHg.

**Conclusion**

As the device has reached the required standards, it is recommended for clinical and personal use in a general population.

**Acknowledgements and Conflict of Interest**

The monitor was supplied for the purposes of the study by the manufacturer OMRON Healthcare CO.,LTD. who also funded the study. None of the authors has any association with OMRON Healthcare CO.,LTD. or has received any personal benefit from OMRON Healthcare Co.,LTD.

## References

1. O'Brien E, Atkins N, Stergiou G, Karpettas N, Parati G, Asmar R, Imai Y, Wang J, Mengden T, Shennan A; on behalf of the Working Group on Blood Pressure Monitoring of the European Society of Hypertension. European Society of Hypertension International Protocol revision 2010 for the Validation of Blood Pressure Measuring Devices In Adults. *Blood Press Monit* 2010;15:23–38.