

**DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE**

A SIGNED COPY WILL BE POSTED ON THE [www.dableducational.org](http://www.dableducational.org) WEBSITE

**SECTION A** - Please complete all items.

I **Kevin Tan,** a Director of **Guangdong Transtek Medical Electronics Co.,Ltd Co.,Ltd ,**

Name of a Company Director

Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker<sup>a</sup> **GuangdongTranstek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**

Manufacturer<sup>b</sup> **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**

Brand<sup>c</sup> **Kinetik Wellbeing** Model<sup>d</sup> **TMB-2080**

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker<sup>a</sup> **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**

Manufacturer<sup>b</sup> **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**

Brand<sup>c</sup> **TRANSTEK** Model<sup>d</sup> **TMB-1491**

Existing validated blood pressure measuring device.

which has previously passed the **2010** protocol, the results of which were published as follows:

Title: Validation of Transtek blood pressure monitor TMB-1491 for self-measurement according to the European Society of Hypertension International Protocol

Authors: Tian H., Zeng S., Zhong X., Gong W. and Liu W.

Publication: Blood Press Monit. 2015 May

Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1–18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>e</sup> <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	

	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>

**An explanation of each item ticked “Yes” must be included in Section B or on a separate sheet.**

- Notes:
- a Provide the name and address of the actual maker of the device.
  - b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
  - c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
  - d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
  - e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
  - f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
  - g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

See attached document

SECTION C Please check that the following are included with the application

- A manual for the validated device [x]
A manual for the device for which equivalence is being sought [x]
Completed DET9 Form [x]
An image of the device for which equivalence is being sought [x]
An image of the screen layout of validated device\* [x]
An image of the screen layout of the device for which equivalence is being sought\* [x]

\* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director Kevin Tan

Company Stamp/Seal

Name Kevin Tan

Date October 14, 2021 Jie Zhu


Signature of Witness

Name Jie.Zhu

Address Zone A, No.105, Dongli Road, Torch Development District, 528437, Zhongshan, Guangdong, China



Comparison of the Kinetik Wellbeing TMB-2080 with the TRANSTEK TMB-1491

Devices – Item 9	Kinetik Wellbeing TMB-2080	TRANSTEK TMB-1491
Pictures		
Display Image		
Validation	/	ESH 2010
Category	<i>Upper Arm device for self measurement of blood pressure</i>	<i>Upper Arm device for self measurement of blood pressure</i>
Casing – Item 10	<p><b>Dimensions</b></p> <p>100x129x45mm</p>	<p><b>Dimensions</b></p> <p>110mm*110mm*40mm</p>

	<p><b>Ports</b></p> <p><i>Cuff port</i></p> <p><b>Features</b></p> <p>Kinetik Wellbeing printing</p> <p>Blood pressure measurement</p> <p>WHO Classification</p> <p>Pulse rate</p> <p>Button printing</p>	<p><b>Ports</b></p> <p><i>Cuff port</i></p> <p><b>Features</b></p> <p>Transtek printing</p> <p>Blood pressure measurement</p> <p>WHO Classification</p> <p>Pulse rate</p> <p>Button printing</p>
<b>Display – Item 11</b>	<p><i>Type</i></p> <p>LCD</p>	<p><i>Type</i></p> <p>LCD</p>
<b>Carrying/Mounting Facilities – Item 12</b>	None	None
<b>Software other than Algorithm – Item 13</b>	<p><i>Dual Users</i></p> <p><i>90 sets memory/per user (2 users total) + guest mode</i></p> <p><i>WHO Indicator</i></p> <p><i>mmHg unit</i></p>	<p><i>One User</i></p> <p><i>60 sets memories/per user</i></p> <p><i>WHO Indicator</i></p> <p><i>mmHg unit or kPa</i></p>
<b>Memory Capacity Item 14</b>	<i>90 sets memories/per user (2 users total) + guest mode</i>	<i>60 sets memories/per user</i>
<b>Printing Facilities Item 15</b>	N/A	N/A
<b>Communication Facilities – Item 16</b>	N/A	N/A

<p><b>Power Supply</b> <b>Item 17</b></p>	<p>4 dry cells 1.5V AAA, 6V DC 5V/1A power adapter</p>	<p>4 dry cells 1.5V AAA, 6V DC</p>
<p><b>Other differences</b></p>	<p>Other Details on Equivalent device that are different to Validated device N/A</p>	<p>Other Details on Validated device that are different to Equivalent device N/A</p>
<p><b>Same Criteria</b></p>	<p><b>Measurement</b></p> <p><b>Accuracy</b> Pressure: 5°C-40°C within ±3mmHg (0.4kPa) Pulse value: ±5%</p> <p><b>Method</b> Oscillographic testing mode</p> <p><b>Ranges</b> Rated cuff pressure: 0 mmHg ~ 299 mmHg Measurement pressure: SYS: 60 mmHg ~ 230 mmHg DIA: 40 mmHg ~ 130 mmHg Pulse value: (40-199) beat/minute</p> <p><b>Inflation</b> Automatic inflation</p> <p><b>Deflation</b> Automatic deflation</p>	<p><b>Measurement</b></p> <p><b>Accuracy</b> Pressure: 5°C-40°C within ±3mmHg (0.4kPa) Pulse value: ±5%</p> <p><b>Method</b> Oscillographic testing mode</p> <p><b>Ranges</b> Rated cuff pressure: 0kpa - 40kpa (0mmHg ~ 300mmHg) Measurement pressure: SYS: 60mmHg ~ 230mmHg (8.0kPa ~ 30.7kPa) DIA: 40mmHg ~ 130mmHg (5.3kPa ~ 17.3kPa) Pulse value: (40-199) beat/minute</p> <p><b>Inflation</b> Automatic inflation</p> <p><b>Deflation</b> Automatic deflation</p>

	<p><b>Cuffs (Please state sizes and materials used)</b></p> <p>22CM-42CM, Nylon</p> <p><b>Sensors</b></p> <p>Piezo-resistive</p> <p><b>Measurement Records</b></p> <p>2 users x 90 readings each (plus guest mode)</p> <p><b>Measurements other than blood pressure</b></p> <p>Pulse rate</p> <p><b>Buttons/Switches</b></p> <p>Power button – ON SYMBOL</p> <p>Memory button - MEM</p> <p>User button – USER 1 &amp; USER 2 SYMBOL</p> <p>Analysis – N/A</p> <p>Event Marking – N/A</p> <p>Communication – N/A</p> <p><b>Display/Symbols/Indicators</b></p> <p><b>Preparation</b></p> <p>Automatic Zero setting</p> <p><b>Measurement Procedure</b></p> <p>Inflation symbol</p> <p>Pressure value indication</p>	<p><b>Cuffs(Please state sizes and materials used)</b></p> <p>22cm-32cm or 22-42cm, nylon</p> <p><b>Sensors</b></p> <p>Piezo-resistive</p> <p><b>Measurement Records</b></p> <p>60 measurement records</p> <p><b>Measurements other than blood pressure</b></p> <p>Pulse rate</p> <p><b>Buttons/Switches</b></p> <p>Power button – START/STOP button</p> <p>Memory button – M button</p> <p>Set button – S button</p> <p>Analysis – N/A</p> <p>Event Marking – N/A</p> <p>Communication – N/A</p> <p><b>Display/Symbols/Indicators</b></p> <p><b>Preparation</b></p> <p>Automatic Zero setting</p> <p><b>Measurement Procedure</b></p> <p>Inflation symbol</p> <p>Pressure value indication</p>
--	---	---

	<p><i>Current time</i></p> <p><b>Post measurement</b></p> <p><i>Upper Arm</i></p> <p><b>Measurement Records</b></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p> <p><i>Measurement time</i></p> <p><i>Memory Query symbol</i></p> <p><b>Date and Time</b></p> <p><i>Time is displayed in the lower left corner of LCD</i></p> <p><b>Power</b></p> <p><i>Low power</i></p> <p><b>Function</b></p> <p><i>Measuring blood pressure and heart rare</i></p> <p><i>Recall measurement records</i></p> <p><i>Delete measurement records</i></p> <p><b>Communication</b></p> <p><i>N/A</i></p> <p><b>Features</b></p> <p><i>Measuring during inflation</i></p> <p><b>Algorithms</b></p> <p><i>Equivalent device has the identical measurement algorithm as the validated device.</i></p>	<p><i>Current time</i></p> <p><b>Post measurement</b></p> <p><i>Upper Arm</i></p> <p><b>Measurement Records</b></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p> <p><i>Measurement time</i></p> <p><i>Memory Query symbol</i></p> <p><b>Date and Time</b></p> <p><i>Time is displayed lower right corner of LCD</i></p> <p><b>Power</b></p> <p><i>Low power</i></p> <p><b>Function</b></p> <p><i>Measuring blood pressure and heart rate</i></p> <p><i>Recall measurement records</i></p> <p><i>Delete measurement records</i></p> <p><b>Communication</b></p> <p><i>N/A</i></p> <p><b>Features</b></p> <p><i>Measuring during inflation</i></p> <p><b>Algorithms</b></p> <p><i>Equivalent device has the identical measurement algorithm as the validated device.</i></p>
--	---	---

<p><b>Comparable Criteria</b></p>	<p><b>Measurement</b></p> <p><i>Cuffs (Please state sizes and materials used)</i></p> <p><i>About 22cm-42cm, Nylon</i></p> <p><b>Measurement Records</b></p> <p><i>90 sets/per user, total two users (plus guest measurement)</i></p> <p><b>Display/Symbols/Indicators</b></p> <p><i>Post Measurement</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p>	<p><b>Measurement</b></p> <p><i>Cuffs (Please state sizes and materials used)</i></p> <p><i>About 22cm-32cm or 22-42cm, nylon</i></p> <p><b>Measurement Records</b></p> <p><i>60 sets/per user, total two users</i></p> <p><b>Display/Symbols/Indicators</b></p> <p><i>Post Measurement</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p>
-----------------------------------	---	---

<p><b>Comments</b></p>	
<p><b>Recommendation</b></p>	<p><b>Recommended</b></p>
<p><b>Date</b></p>	<p><b>January 2022</b></p>