

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2013

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I Willis Chan,  
Name of a Company Director

a Director of Microlife Corporation Co.,,  
Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker<sup>a</sup> Microlife Corporation Co., Address 9F,431,RuiGuang Road,Nei-Hu,Taipei,114,Taiwan,R.O.C  
 Manufacturer<sup>b</sup> KAZ Home Appliances Address Flat 4B&4C,Productivity Building,2nd High Technology Road,Science and industry Park, NanShan District,Shenzhen,PRC  
 Brand<sup>c</sup> BRAUN Model<sup>d</sup> BP2000

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker<sup>a</sup> Microlife Corporation Co., Address 9F,431,RuiGuang Road,Nei-Hu,Taipei,114,Taiwan,R.O.C  
 Manufacturer<sup>b</sup> Microlife Corporation Co., Address 9F,431,RuiGuang Road,Nei-Hu,Taipei,114,Taiwan,R.O.C  
 Brand<sup>c</sup> Microlife Model<sup>d</sup> W100

Existing validated blood pressure measuring device.

which has previously passed the ESH, BHS protocol, the results of which were published as follows:

Paolo Palatini, Francesca Dorigatti, Elisa Bonso and Fabio Ragazzo; Validation of Microlife BP W100 wrist device assessed according to the European Society of Hypertension and the British Hypertension Society protocols.

Full reference

Blood Press Monit 2009, 14:41-44

The only differences between the devices involve the following components:

Tick one box for each item 1-18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>e</sup> <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	18	Other Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

- Notes:
- a Provide the name and address of the actual maker of the device.
  - b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
  - c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
  - d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
  - e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
  - f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
  - g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.


Attachment: BP2000 Comparison table for more explanation on items No 9, 10, 11, 14 and 18.

SECTION C Please check that the following are included with the application

- A manual for the validated device
- A manual for the device for which equivalence is being sought
- An image of the validated device
- An image of the device for which equivalence is being sought
- An image of the screen layout of validated device\*
- An image of the screen layout of the device for which equivalence is being sought\*

\* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director  Company Stamp/Seal


Name

Willis Chan

Date

Jan 28, 2014

Signature of Witness



Name

Gerhard Frick



Address

Microlife AG, 9443 Widnau, Switzerland

**microlife**

Microlife AG  
Epenstrasse 139  
9443 Widnau / Switzerland  
Phone +41 / 71 727 70 30  
Fax +41 / 71 727 70 39

### Comparison of the BRAUN BP2000 with the Microlife BP W100

Devices	BRAUN BP2000	9	Microlife W100	9
Image		10		10
Validation			ESH, BHS	
LCD Display		11		11
Device Criteria	<b>Memory Capacity for stored values:</b> - 10 sets - shown with symbol «M»	14	<b>Memory Capacity for stored values:</b> - 200 sets - shown with symbol «M» and date and time	14
	<b>Other Facilities:</b> <b>Display/Symbols/Indicators</b> - Irregular Heartbeat Indicator: NO - Pulse Beep during measurement: NO - Date and Time: NO - Measurement range (heart rate): 0 - 199	18	<b>Other Facilities:</b> <b>Display/Symbols/Indicators</b> - Irregular Heartbeat Indicator: YES - Pulse Beep during measurement: YES - Date and Time: YES (2 alarm times i.e. for medication) - Measurement range (heart rate): 0 - 200	18
Web Link			<a href="http://www.microlife.com/products/hypertension/wrist/bp-w100/">http://www.microlife.com/products/hypertension/wrist/bp-w100/</a>	

Comparison of the Braun BP 2000 (BBP2000) with the Microlife BP W100

Devices	Braun BP 2000 (BBP2000)	Microlife BP W100
Pictures		
Display		
Validation		ESH
Device 1 Criteria		
Same Criteria	<p><b>Measurement</b></p> <p><i>Accuracy</i></p> <p>BP accuracy ± 3 mmHg 1, 5</p> <p>Pulse accuracy ± 5% 1, 5</p> <p><i>Method</i></p> <p>Oscillometric measurement method 1, 5</p> <p>BP 20 mmHg – 280 mmHg 1, 5, 7, 8</p> <p>Manually initiated measurements 13</p> <p>Measurements are from single inflations 13</p> <p><i>Inflation</i></p> <p>Inflation 0 mmHg – 300 mmHg 1, 5, 7</p> <p>Automatic Inflation 7</p>	<p><b>Measurement</b></p> <p><i>Accuracy</i></p> <p>BP accuracy ± 3 mmHg 1, 5</p> <p>Pulse accuracy ± 5% 1, 5</p> <p><i>Method</i></p> <p>Oscillometric measurement method 1, 5</p> <p>BP 20 mmHg – 280 mmHg 1, 5, 7, 8</p> <p>Manually initiated measurements 13</p> <p>Measurements are from single inflations 13</p> <p><i>Inflation</i></p> <p>Inflation 0 mmHg – 300 mmHg 1, 5, 7</p> <p>Automatic Inflation 7</p>

Devices	Braun BP 2000 (BBP2000)	Microlife BP W100
<b>Same Criteria (Continued)</b>	<p><b>Measurement (Continued)</b></p> <p><i>Deflation</i></p> <p>Automatic Deflation 8</p> <p><i>Cuffs</i></p> <p>Wrist circ. 13.5 cm – 21.5 cm 6</p> <p><i>Sensors</i></p> <p>Pressure sensor: capacitive 5</p> <p><b>Buttons/Switches</b></p> <p><i>Power</i></p> <p>On/Off with Start/Stop (Ⓢ symbol) 10</p> <p><i>Measurement Records</i></p> <p>Memory 10</p> <p><b>Display/Symbols/Indicators</b></p> <p><i>Measurement Procedure</i></p> <p>During Measurement: BP Level &amp; Heartbeat 11</p> <p>Beep after measurement 18</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>Measurement error Err 1, Err 2, Err 3, Err 5, Hi, Lo 11</p> <p><i>Measurement Records</i></p> <p>Memory “M” symbol 11</p> <p>Memory recall number (Replaces pulse rate momentarily) 11</p> <p><i>Power</i></p> <p>Low and exhausted battery 11, 17</p> <p><b>Casing</b></p> <p><i>Display</i></p> <p>Single screen display 10</p> <p>Segment LCD 10</p> <p><i>Power</i></p> <p>2 “AAA” batteries ~ 200 measurements<sup>2</sup> 17</p> <p>Automatic switch-off when not used for 1 min 17</p> <p>Rechargeable batteries permitted but not recommended<sup>3</sup> 17</p>	<p><b>Measurement (Continued)</b></p> <p><i>Deflation</i></p> <p>Automatic Deflation 8</p> <p><i>Cuffs</i></p> <p>Wrist circ. 13.5 cm – 21.5 cm 6</p> <p><i>Sensors</i></p> <p>Pressure sensor: capacitive 5</p> <p><b>Buttons/Switches</b></p> <p><i>Power</i></p> <p>On/Off with Start/Stop (Ⓢ symbol) 10</p> <p><i>Measurement Records</i></p> <p>Memory 10</p> <p><b>Display/Symbols/Indicators</b></p> <p><i>Measurement Procedure</i></p> <p>During Measurement: BP Level &amp; Heartbeat 11</p> <p>Beep after measurement 18</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>Measurement error Err 1, Err 2, Err 3, Err 5, Hi, Lo 11</p> <p><i>Measurement Records</i></p> <p>Memory “M” symbol 11</p> <p>Memory recall number (Replaces pulse rate momentarily) 11</p> <p><i>Power</i></p> <p>Low and exhausted battery 11, 17</p> <p><b>Casing</b></p> <p><i>Display</i></p> <p>Single screen display 10</p> <p>Segment LCD 10</p> <p><i>Power</i></p> <p>2 “AAA” batteries 17</p> <p>Automatic switch-off when not used for 1 min 17</p> <p>Rechargeable batteries permitted 17</p>
<b>Comparable Criteria</b>	<p><b>Measurement</b></p> <p><i>Method</i></p> <p>Pulse 40 bpm – 199 bpm<sup>1</sup> 1, 5, 8</p> <p><i>Measurement Records</i></p> <p>Memory: 10 measurements 14</p>	<p><b>Measurement</b></p> <p><i>Method</i></p> <p>Pulse 40 bpm – 200 bpm<sup>1</sup> 1, 5, 8</p> <p><i>Measurement Records</i></p> <p>Memory: 200 measurements 14</p>

Devices	Braun BP 2000 (BBP2000)	Microlife BP W100
Device 2 Criteria		<p><b>Buttons/Switches</b></p> <p><i>Settings</i></p> <p>Date/Time set 10</p> <p><b>Display/Symbols/Indicators</b></p> <p><i>Measurement Procedure</i></p> <p>Audible pulse indicator during deflation 18</p> <p><i>Post Measurement</i></p> <p>Irregular heartbeat 11, 13, 18</p> <p><i>Date and Time</i></p> <p>Date and Time 11</p> <p>Date and Time (During memory recall) 11</p> <p>Alarm reminder (2 alarms/day) / Alarm reminder off 18</p> <p><b>Algorithms</b></p> <p><i>Diagnostic</i></p> <p>Irregular heartbeat detection 13</p>

<b>Comments</b>	Microlife are acting as an OEM manufacturer for the Braun BP2000. The device is essentially a more basic version of the BP W100.	
	1	<p><b>Query</b> The manual for the Braun BP2000 (and BP2200) states that the measurement range for the pulse is 40 bpm to-199 bpm whereas the manual for the Microlife BP W100 states that the measurement range for the pulse is 40 bpm to 200 bpm. Why the slight difference?</p> <p><b>Response</b> It is because of the display limitation. The 3rd digit in the Braun is only a “1”, whereas, in the Microlife, it can show a “2”. We have listed the difference in the comparison table.</p> <p><b>Comment</b> The explanation is accepted.</p>
		<p><b>Query</b> The manual for the Braun BP2000 (and BP2200) states that new batteries will provide approximately 200 measurements. No such information is provided in the manual for the Microlife BP W100. How many measurements can be made by the Microlife BP W100 with a fresh set of batteries?</p> <p><b>Response</b> Both, the Braun and the Microlife devices have the same power supply and electronic circuitry, which is good for 400 measurement cycles with usual standard quality batteries. For the avoidance of enquiries and user complaints in case of lower quality batteries, Braun prefers to claim only 200 measurements in their user manual. This does not mean any technical difference between the products.</p> <p><b>Comment</b> The explanation is accepted.</p>
		<p><b>Query</b> The manual for the Microlife BP W100 states that rechargeable batteries are permitted. The manual for the Braun BP2000 (and BP2200) contains no information on whether or not rechargeable batteries are permitted. Are rechargeable batteries permitted for the Braun BP2000?</p> <p><b>Response</b> Both, the Braun and the Microlife devices have the same power supply and electronic circuitry which also allows the use of standard quality rechargeable batteries. For the avoidance of enquiries or complaints due to use of poor quality or exhausted rechargeable batteries, Braun prefers to abstain from recommending rechargeable batteries at all. This does not mean any technical difference between the products.</p> <p><b>Comment</b> The explanation is accepted.</p>
<b>Recommendation</b>	Equivalence is Recommended	
<b>Date</b>	13 February 2014	