

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2013

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I **Gerhard Frick,** a Director of **Microlife AG,**  
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker<sup>a</sup> **ONBO** Address **497 Dalang South Road, Longhua, Shenzhen, Guangdong, China**

Manufacturer<sup>b</sup> **Microlife AG** Address **Espenstrasse 139, 9444 Widnau**

Brand<sup>c</sup> **Microlife** Model<sup>d</sup> **A2 Classic / BP 3UG1-2E**

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker<sup>a</sup> **ONBO** Address **497 Dalang South Road, Longhua, Shenzhen, Guangdong, China**

Manufacturer<sup>b</sup> **Microlife AG** Address **Espenstrasse 139, 9444 Widnau**

Brand<sup>c</sup> **Microlife** Model<sup>d</sup> **BP 3BT0-A**

Existing validated blood pressure measuring device.

which has previously passed the BHS protocol, the results of which were published as follows:

Reinders A, Cuckson AC, Lee JTM, Shennan AH. An accurate automated blood pressure device for use in pregnancy and pre-eclampsia: the Microlife 3BT0-A. BJOG 2005;112(7):915-920

Refer to attached documents.

Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1-18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>e</sup> <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>g</sup> <input type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>g</sup> <input type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	18	Other Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

- Notes:
- a Provide the name and address of the actual maker of the device.
  - b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
  - c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
  - d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
  - e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
  - f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
  - g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

**SECTION B** An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.


As attached file : A2 Basic Comparison items No 9, 10, 11, 14, 18 are explained in the attached table.

**SECTION C** Please check that the following are included with the application

- A manual for the validated device
- A manual for the device for which equivalence is being sought
- An image of the validated device
- An image of the device for which equivalence is being sought
- An image of the screen layout of validated device\*
- An image of the screen layout of the device for which equivalence is being sought\*

\* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

**SECTION D** Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to [info@dableducational.org](mailto:info@dableducational.org).

Signature of Director  Company Stamp/Seal

Name Gerhard Frick

Date 2016-04-08

Signature of Witness 

Name Jerry Lin

Address 9F, NO.431, RuiGuang Road, Nei-Hu,  
Taipei, 11492, Taiwan. R.O.C

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2013

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I **Gerhard Frick,** a Director of **Microlife AG,**  
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker<sup>a</sup> **ONBO** Address **497 Dalang South Road, Longhua, Shenzhen, Guangdong, China**

Manufacturer<sup>b</sup> **Microlife AG** Address **Espenstrasse 139, 9444 Widnau**

Brand<sup>c</sup> **Microlife** Model<sup>d</sup> **A2 Classic / BP 3UG1-2E**

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker<sup>a</sup> **ONBO** Address **497 Dalang South Road, Longhua, Shenzhen, Guangdong, China**

Manufacturer<sup>b</sup> **Microlife AG** Address **Espenstrasse 139, 9444 Widnau**

Brand<sup>c</sup> **Microlife** Model<sup>d</sup> **BP A100**

Existing validated blood pressure measuring device.

which has previously passed the ESH protocol, the results of which were published as follows:

Bonso E, Dorigatti F, Palatini P. Accuracy of the BP A100 blood pressure measuring device coupled with a single cuff with standard-size bladder over a wide range of arm circumferences. Blood Press Monit 2009;14:216-9

Refer to attached documents.

Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1-18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>e</sup> <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>g</sup> <input type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>g</sup> <input type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	18	Other Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

- Notes:
- a Provide the name and address of the actual maker of the device.
  - b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
  - c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
  - d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
  - e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
  - f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
  - g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.




As attached file : A2 Basic Comparison items No 9, 10, 11, 14, 18 are explained in the attached table.

SECTION C Please check that the following are included with the application
A manual for the validated device [X]
A manual for the device for which equivalence is being sought [X]
An image of the validated device [X]
An image of the device for which equivalence is being sought [X]
An image of the screen layout of validated device\* [X]
An image of the screen layout of the device for which equivalence is being sought\* [X]
\* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director [Signature] Company Stamp/Seal
Name Gerhard Frick
Date 2016-04-08
Signature of Witness [Signature]
Name Jerry Lin
Address 9F,NO.431,RuiGuang Road,Nei-Hu, Taipei,11492,Taiwan.R.O.C

Comparison of the Microlife A2 Classic with the Microlife 3BT0-A and Microlife BP A100

Devices	Microlife A2 Classic (BP 3UG1-2E)	9	Microlife BP 3BT0-A	9	Microlife BP A100	9
Image		10		10		10
Validation			BHS		ESH 2002	
LCD Display		11		11		11
	<b>Memory Capacity for stored values:</b> - 30 sets - shown with symbol «M» - allows indicate all-memory average (see I/B)	14	<b>Memory Capacity for stored values:</b> - 1 set - shown with symbol «M»	14	<b>Memory Capacity for stored values:</b> - 1 set - shown with symbol «M»	14
	<b>Other Facilities:</b> <b>Display/Symbols/Indicators</b>	18	<b>Other Facilities</b> <b>Display/Symbols/Indicators</b>	18	<b>Other Facilities</b> <b>Display/Symbols/Indicators</b>	18

	<p><b>Pulse Arrhythmia Indicator (PAD):</b> Yes (indicates pulse irregularities during measurement which may affect the reading)</p> <p><b>Pulse Beep during measurement:</b> No (less disturbance for the patient)</p> <p><b>Cuff compartment:</b> No</p> <p><b>Measurement</b> <i>Accuracy</i> Blood Pressure Accuracy <math>\pm</math> 3 mmHg Pulse Accuracy <math>\pm</math> 5%</p> <p><i>Method</i> Oscillometric</p> <p><i>Ranges</i> Cuff pressure: 0 -299 mmHg Measurement: 20 mmHg – 280 mmHg</p> <p>Pulse rate: 40-200 beats/minute</p> <p><b>Traffic Light Indicator:</b> Yes (following WHO 2003)</p> <p><b>Cuffs:</b> Microlife M-Cuff (22-32cm) <sup>1)</sup> Microlife L-Cuff (32-42cm) <sup>1)</sup> Microlife M-L-Cuff (22-42cm) <sup>2)</sup> Microlife M-L-Rigid Conical Cuff (22-42cm) <sup>3)</sup></p>	<p><b>Pulse Arrhythmia Indicator (PAD):</b> No</p> <p><b>Pulse Beep during measurement:</b> Yes</p> <p><b>Cuff compartment:</b> No</p> <p><b>Measurement</b> <i>Accuracy</i> Blood Pressure Accuracy <math>\pm</math> 3 mmHg Pulse Accuracy <math>\pm</math> 5%</p> <p><i>Method</i> Oscillometric</p> <p><i>Ranges</i> Cuff pressure: 0 -299 mmHg Measurement: 30 mmHg – 280 mmHg (no separate range for SBP and DBP specified) Pulse rate: 40-200 beats/minute</p> <p><b>Traffic Light Indicator:</b> No</p> <p><b>Cuffs:</b> Microlife AC-1-M-Cuff (22-32cm) <sup>1)</sup> Microlife AC-1-L-Cuff (32-42cm) <sup>1)</sup></p>	<p><b>Pulse Arrhythmia Indicator (PAD):</b> Yes (indicates pulse irregularities during measurement which may affect the reading)</p> <p><b>Pulse Beep during measurement:</b> Yes</p> <p><b>Cuff compartment:</b> Yes</p> <p><b>Measurement</b> <i>Accuracy</i> Blood Pressure Accuracy <math>\pm</math> 3 mmHg Pulse Accuracy <math>\pm</math> 5%</p> <p><i>Method</i> Oscillometric</p> <p><i>Ranges</i> Cuff pressure: 0 -299 mmHg Measurement: 20 mmHg – 280 mmHg (no separate range for SBP and DBP specified) Pulse rate: 40-200 beats/minute</p> <p><b>Traffic Light Indicator:</b> No</p> <p><b>Cuffs:</b> Microlife M-Cuff (22-32cm) <sup>1)</sup> Microlife L-Cuff (32-42cm) <sup>1)</sup> Microlife M-L-Cuff (22-42cm) <sup>2)</sup> Microlife M-L-Rigid Conical Cuff (22-42cm) <sup>3)</sup></p>
<p><b>Reference documents</b></p>	<p><sup>1)</sup> Reference dev. BP 3BTO-A – validated with standard Microlife AC-1-L-Cuff and AC-1-M-Cuff Cuckson AC, Reinders A, Shabeeh H, Shennan AH. Validation of the Microlife BP 3BTO-A oscillometric blood pressure monitoring device according to a modified British Hypertension Society protocol Blood Press Monit 2002;7(6):319-324</p>		

	<p><sup>2)</sup> Reference dev. BP A100 – validated with Microlife M-L-Cuff (22-42cm)  <i>Bonso E, Dorigatti F, Palatini P. Accuracy of the BP A100 blood pressure measuring device coupled with a single cuff with standard-size bladder over a wide range of arm circumferences. Blood Press Monit 2009;14:216-9</i></p> <p><sup>3)</sup> Reference dev. BP A100 – validated with Microlife M-L-Cuff Rigid Conical Cuff (22-42cm)  <i>Bonso E, Saladini F, Zanier A, Benetti E, Dorigatti F, Palatini P. Accuracy of a single rigid conical cuff with standard-size bladder coupled to an automatic oscillometric device over a wide range of arm circumferences. Hypertens Res. 2010;33(11):1186-91</i></p>		
<b>Web link</b>	n/a	<a href="http://www.microlife.com/products/hypertension/automatic/bp-3bt0-a-2/">http://www.microlife.com/products/hypertension/automatic/bp-3bt0-a-2/</a>	n/a
<b>Recommendation</b>	<b>Equivalence is recommended</b>		
<b>Date</b>	<b>9 May 2016</b>		