

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2013

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I **Kevin Tan,** a Director of **Guangdong Transtek Medical Electronics Co.,Ltd**,
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker^a **Veridian Healthcare, LLC** Address **1175 Lakeside Drive, Gurnee, Illinois 60031**
 Manufacturer^b **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**
 Brand^c **SmartHeart** Model^d **01-509**

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker^a **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**
 Manufacturer^b **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**
 Brand^c **TRANSTEK** Model^d **LS808-B**

Existing validated blood pressure measuring device.

which has previously passed the **ESH 2010** protocol, the results of which were published as follows:

Title: Validation of Transtek LS808-B for self/home measurement according to the European Society of Hypertension International Protocol revision 2010

Authors: Zhong Hua Liu

Publication: Blood Pressure Monitoring 2016 Dec; 21(6):352-55

Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1–18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^e <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^e <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^e <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^e <input checked="" type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

Notes: a Provide the name and address of the actual maker of the device.

b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.

- c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
- d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
- e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
- f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
- g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

[See attached document](#)

SECTION C Please check that the following are included with the application

- A manual for the validated device
- A manual for the device for which equivalence is being sought
- An image of the validated device
- An image of the device for which equivalence is being sought
- An image of the screen layout of validated device*
- An image of the screen layout of the device for which equivalence is being sought*

* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director Kevin Tan

Name Kevin Tan

Date January 11st, 2018

Signature of Witness Havana Hu

Name Havana Hu


Address Zone A, No.105 ,Dongli Road, Torch Development District,

Zhongshan, 528437, Guangdong, China



Company Stamp/Seal

Comparison of the SmartHeart 01-509 with the TRANSTEK LS808-B

Devices – Item 9	SmartHeart 01-509	TRANSTEK LS808-B
Pictures		
Display Image		
Validation	Upper arm device for self measurement of blood pressure	ESH 2010
Category	Upper arm device for self measurement of blood pressure	Upper arm device for self measurement of blood pressure
Casing – Item 10	Dimensions 130.9mm*73mm*29.4mm	Dimensions 130.9mm*73mm*29.4mm

	<p><i>Ports</i></p> <p><i>Cuff port and DC power port</i></p> <p><i>Features</i></p> <p>Cuff and AC adaptor connectors</p> <p>Model name printing</p> <p>Button printing</p> <p>SYS, DIA, Pul/min printing</p>	<p><i>Ports</i></p> <p><i>Cuff port and DC power port</i></p> <p><i>Features</i></p> <p>Cuff and AC adaptor connectors</p> <p>Model name printing</p> <p>Button printing</p> <p>SYS, DIA, Pul/min printing</p>
Display – Item 11	<i>LCD</i>	<i>LCD</i>
Carrying/Mounting Facilities – Item 12	None	<i>None</i>
Software other than Algorithm – Item 13	<p>Dual Users</p> <p>250 sets memories/per user</p> <p>AHA indicator</p> <p>mmHg unit</p>	<p>Dual Users</p> <p>60 sets memories/per user</p> <p>WHO indicator</p> <p>mmHg unit</p>
Memory Capacity Item 14	250 sets memories/per user	60 sets memories/per user
Printing Facilities	N/A	N/A

Item 15		
Communication Facilities – Item 16	N/A	N/A
Power Supply Item 17	1. lithium battery 2. 6V DC Jack	1. lithium battery 2. 6V DC Jack
Other differences	<i>Other Details on Equivalent device that are different to Validated device</i> N/A	<i>Other Details on Validated device that are different to Equivalent device</i> N/A
Same Criteria	<p>Measurement</p> <p><i>Accuracy</i></p> <p><i>Pressure:</i></p> <p>5°C-40°C within±3mmHg(0.4kPa)</p> <p><i>Pulse value:±5%</i></p> <p><i>Method</i></p> <p><i>Oscillographic testing mode</i></p> <p><i>Ranges</i></p> <p><i>Rated cuff pressure:</i></p> <p>0mmHg~299mmHg(0kPa ~ 39.9kPa)</p> <p><i>Measurement pressure:</i></p> <p>SYS: 60mmHg~230mmHg (8.0kPa~30.7kPa)</p>	<p>Measurement</p> <p><i>Accuracy</i></p> <p><i>Pressure:</i></p> <p>5°C-40°C within±3mmHg(0.4kPa)</p> <p><i>Pulse value:±5%</i></p> <p><i>Method</i></p> <p><i>Oscillographic testing mode</i></p> <p><i>Ranges</i></p> <p><i>Rated cuff pressure:</i></p> <p>0mmHg~299mmHg(0kPa ~ 39.9kPa)</p> <p><i>Measurement pressure:</i></p> <p>SYS: 60mmHg~230mmHg (8.0kPa~30.7kPa)</p>

	<p><i>DIA: 40mmHg~130mmHg (5.3kPa~17.3kPa)</i></p> <p><i>Pulse value: (40-199)beat/minute</i></p> <p><i>Inflation</i></p> <p><i>Automatic inflation</i></p> <p><i>Deflation</i></p> <p><i>Automatic deflation</i></p> <p><i>Sensors</i></p> <p><i>Piezo-resistive</i></p> <p><i>Measurements other than Blood Pressure</i></p> <p><i>Pluse rate</i></p> <p>Buttons/Switches</p> <p><i>Power</i></p> <p><i>User 1 / User 2 botton</i></p> <p><i>Measurement Records</i></p> <p><i>MEM button</i></p>	<p><i>DIA: 40mmHg~130mmHg (5.3kPa~17.3kPa)</i></p> <p><i>Pulse value: (40-199)beat/minute</i></p> <p><i>Inflation</i></p> <p><i>Automatic inflation</i></p> <p><i>Deflation</i></p> <p><i>Automatic deflation</i></p> <p><i>Sensors</i></p> <p><i>Piezo-resistive</i></p> <p><i>Measurements other than Blood Pressure</i></p> <p><i>Pluse rate</i></p> <p>Buttons/Switches</p> <p><i>Power</i></p> <p><i>User 1 / User 2 botton</i></p> <p><i>Measurement Records</i></p> <p><i>MEM button</i></p>
--	---	---

	<p><i>Function</i></p> <p><i>User 1 / User 2/ MEM/“lock” button</i></p> <p><i>Display/Symbols/Indicators</i></p> <p><i>Preparation</i></p> <p><i>Automatic Zero setting</i></p> <p><i>Measurement Procedure</i></p> <p><i>Inflation symbol</i></p> <p><i>Pressure value indication</i></p> <p><i>Current time</i></p> <p><i>Measurement Records</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p> <p><i>Measurement time</i></p> <p><i>Memory Query symbol</i></p> <p><i>Power</i></p> <p><i>Low power</i></p>	<p><i>Function</i></p> <p><i>User 1 / User 2/ MEM/“lock” button</i></p> <p><i>Display/Symbols/Indicators</i></p> <p><i>Preparation</i></p> <p><i>Automatic Zero setting</i></p> <p><i>Measurement Procedure</i></p> <p><i>Inflation symbol</i></p> <p><i>Pressure value indication</i></p> <p><i>Current time</i></p> <p><i>Measurement Records</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p> <p><i>Measurement time</i></p> <p><i>Memory Query symbol</i></p> <p><i>Power</i></p> <p><i>Low power</i></p>
--	---	---

	<p><i>Features</i></p> <p><i>Measuring during inflation</i></p> <p>Algorithms</p> <p><i>Equivalent device has the identical measurement algorithm as the validated device.</i></p>	<p><i>Features</i></p> <p><i>Measuring during inflation</i></p> <p>Algorithms</p> <p><i>Equivalent device has the identical measurement algorithm as the validated device.</i></p>
<p>Comparable Criteria</p>	<p>Measurement</p> <p><i>Cuffs (Please state sizes and materials used)</i></p> <p><i>About 22cm-32cm or 22cm-42cm,polyester</i></p> <p><i>Measurement Records</i></p> <p><i>250 sets/per user,totel two users</i></p> <p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p> <p><i>Measurement time</i></p>	<p>Measurement</p> <p><i>Cuffs (Please state sizes and materials used)</i></p> <p><i>About 22cm-42cm,polyester</i></p> <p><i>Measurement Records</i></p> <p><i>60 sets/per user,totel two users</i></p> <p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p> <p><i>Measurement time</i></p>

	<p><i>Function</i></p> <p><i>Measure blood pressure and heart rate</i></p> <p><i>Recall measurement records</i></p> <p><i>Delete measurement records</i></p>	<p><i>Transmit symbol</i></p> <p><i>Function</i></p> <p><i>Measure blood pressure and heart rate</i></p> <p><i>Recall measurement records</i></p> <p><i>Delete measurement records</i></p> <p><i>Transmit measurement record to APP</i></p>
--	--	---

Comments	
Recommendation	Recommended
Date	19 January 2018