

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2013

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I **Kevin Tan,** a Director of **Guangdong Transtek Medical Electronics Co.,Ltd**,
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker^a **Kaz Europe Sàrl** Address **Place Chauderon 18, 1003 Lausanne, Switzerland**
 Manufacturer^b **Transtek** Address
 Brand^c **Braun** Model^d **BUA6150**

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker^a **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**
 Manufacturer^b **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**
 Brand^c **TRANSTEK** Model^d **TMB-986**

Existing validated blood pressure measuring device.

which has previously passed the **2002** protocol, the results of which were published as follows:

Title: Validation of the TRANSTEK blood pressure monitor TMB-986 for home blood pressure monitoring according to the International Protocol

Authors: Wen Jun Liu; Su Gang Li; Zhe Song; Wei Gong

Publication: Blood Pressure Monitoring. 15(5):278-280, OCT 2010

Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1–18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^e <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	13	Software other than Algorithm	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

- Notes: a Provide the name and address of the actual maker of the device.
 b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
 c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
 d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
 e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.

g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

[See attached document](#)

SECTION C Please check that the following are included with the application

- A manual for the validated device
- A manual for the device for which equivalence is being sought
- An image of the validated device
- An image of the device for which equivalence is being sought
- An image of the screen layout of validated device*
- An image of the screen layout of the device for which equivalence is being sought*

* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director Kevin Tan  Company Stamp/Seal

Name Kevin Tan



Date September 6, 2018

Signature of Witness Wan Hu

Name Wan Hu

Address Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China

Comparison of the BRAUN BUA6150 with the TRANSTEK TMB-986

Devices – Item 9	BRAUN BUA6150	TRANSTEK TMB-986
Pictures		
Display Image		
Validation		ESH 2010 ESH 2002 BHS AAMI
Category	Upper arm device for self-measurement of blood pressure	Upper arm device for self-measurement of blood pressure
Casing – Item 10	<p><i>Dimensions</i> 110mm*124mm*113mm</p> <p><i>Ports</i> Cuff port</p> <p><i>Features</i></p>	<p><i>Dimensions</i> 182mm*100mm*39mm</p> <p><i>Ports</i> Cuff port and DC power port</p> <p><i>Features</i></p>

	User buttons: start/stop, user selection (slide switch), average button, date & time setting buttons Cuff port Display with WHO colour indicator Branding logo and function printing on buttons Battery compartment	User buttons: Start/stop, SET & MEM, User Select Cuff and DC adaptor connectors Model name printing & brand logo Display Battery compartment
Display – Item 11	<i>Type</i> LCD (negative type – white on black background)	<i>Type</i> LCD (black on white background)
Carrying/Mounting Facilities – Item 12	N/A	N/A
Software other than Algorithm – Item 13	40 sets memories/per user (2*40) 4 grade indicator mmHg unit	40 sets memories/per user (2*60) 6 grade indicator mmHg unit
Memory Capacity Item 14	40 sets memories/per user	60 sets memories/per user
Printing Facilities Item 15	N/A	N/A
Communication Facilities – Item 16	N/A	N/A
Power Supply Item 17	4*AA batteries	4*AAA batteries DC power socket (6V)
Other differences	N/A	N/A
Same Criteria	Measurement <i>Accuracy</i> Pulse accuracy $\pm 5\%$ <i>Method</i> Oscillometric measurement method Manually initiated measurement Measurements are from single inflations <i>Ranges</i>	Measurement <i>Accuracy</i> Pulse accuracy $\pm 5\%$ <i>Method</i> Oscillometric measurement method Manually initiated measurement Measurements are from single inflations <i>Ranges</i>

	<p>Rated cuff pressure: 0 mmHg – 300 mmHg Pulse 40 bpm – 199 bpm</p> <p><i>Inflation</i> Automatic inflation by internal pump Zero pressure check before inflation</p> <p><i>Deflation</i> Automatic Deflation Automatic safety release</p> <p><i>Cuffs (Please state sizes and materials used)</i> Nylon Small/Medium (Arm circ. 22 cm to 32 cm) # TMB-1250-02 Large/XLarge (Arm circ. 32-42 cm) # TMB-1250-03</p> <p><i>Sensors</i> Piezo-resistive (semiconductor) pressure sensor</p> <p><i>Measurement Records</i> YES: SYS, DIA, Pulse, IHB, Date & time</p> <p><i>Measurements other than Blood Pressure</i> Pulse (heart rate) Irregular heartbeat Date & time</p> <p>Buttons/Switches <i>Power</i> On/Off with Start/Stop (Start Label)</p> <p><i>Measurement Records</i> Memory User ID (A or B)</p> <p>Display/Symbols/Indicators <i>Preparation</i> Zero pressure adjust - arrow down symbol</p>	<p>Rated cuff pressure: 0 mmHg – 300 mmHg Pulse 40 bpm – 199 bpm</p> <p><i>Inflation</i> Automatic inflation by internal pump Zero pressure check before inflation</p> <p><i>Deflation</i> Automatic Deflation Automatic safety release</p> <p><i>Cuffs (Please state sizes and materials used)</i> Nylon Small/Medium (Arm circ. 22 cm to 32 cm) # AC2232-01 Large/XLarge (Arm circ. 32-42 cm) # TMB-986-AC-05</p> <p><i>Sensors</i> Piezo-resistive (semiconductor) pressure sensor</p> <p><i>Measurement Records</i> YES: SYS, DIA, Pulse, IHB, Date & time</p> <p><i>Measurements other than Blood Pressure</i> Pulse (heart rate) Irregular heartbeat Date & time</p> <p>Buttons/Switches <i>Power</i> On/Off with Start/Stop (Start/Stop Label)</p> <p><i>Measurement Records</i> Memory User ID (A or B)</p> <p>Display/Symbols/Indicators <i>Preparation</i> Zero pressure adjust - arrow down symbol</p>
--	--	---

	<p><i>Measurement Procedure</i> During Measurement: cuff pressure level & heartbeat symbol</p> <p><i>Post Measurement</i> SBP, DBP and Pulse BP classification (WHO)</p> <p><i>Measurement Records</i> Memory recall number User (A or B)</p> <p><i>Date and Time</i> Date and Time (During memory recall)</p> <p><i>Power</i> Low battery indicator</p> <p><i>Function</i> N/A</p> <p><i>Communication</i> N/A</p> <p><i>Features</i> average of last 3 records</p> <p><i>Communication</i> N/A</p>	<p><i>Measurement Procedure</i> During Measurement: cuff pressure level & heartbeat symbol</p> <p><i>Post Measurement</i> SBP, DBP and Pulse BP classification (WHO)</p> <p><i>Measurement Records</i> Memory recall number User (A or B)</p> <p><i>Date and Time</i> Date and Time (During memory recall)</p> <p><i>Power</i> Low battery indicator</p> <p><i>Function</i> N/A</p> <p><i>Communication</i> N/A</p> <p><i>Features</i> average of last 3 records</p> <p><i>Communication</i> N/A</p>
<p>Comparable Criteria</p>	<p>Measurement <i>Accuracy</i> BP accuracy ± 3 mmHg (10°C-40°C)</p> <p><i>Measurement Records</i> Memory: 40 measurements × 2 users</p> <p>Buttons/Switches <i>Settings</i> Date/Time set</p>	<p>Measurement <i>Accuracy</i> BP accuracy ± 3 mmHg (15°C-25°C) ± 6 mmHg otherwise</p> <p><i>Measurement Records</i> Memory: 60 measurements × 2 users</p> <p>Buttons/Switches <i>Settings</i> Set</p>

	<p>Display/Symbols/Indicators <i>Post Measurement</i> Measurement error E1, E2, E3, E4, Eexx Hypertension (Indicator strip) Average (Icon)</p> <p><i>Measurement Records</i> Memory “M” symbol</p> <p><i>Date and Time</i> Date and Time</p> <p>Casing <i>Power</i> 4 “AA” batteries ~ 300 measurements</p>	<p>Display/Symbols/Indicators <i>Post Measurement</i> Measurement error E1, E2, E3, (E10, E11) → E4, E20, E21, Eexx Hypertension (Grading strip) Average (AVG)</p> <p><i>Measurement Records</i> Memory icon (Magnifying glass)</p> <p><i>Date and Time</i> Setting of Date and Time set but only display of Time</p> <p>Casing <i>Power</i> 4 “AAA” batteries</p>
--	---	--

Comments		<p>Braun BUA6150 is similar to previous Braun BP6000 series (BP600, BP6100, BP6200)</p> <p>Braun BP6000 is already ESH approved by equivalence to Transtek TMB-986</p>
Recommendation	Recommended	
Date	15 November 2018	