

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I Yasuhiko Shinozaki, a Director of A&D Compnay,Limited,
 Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker^a A&D Compnay,Limited Address 3-23-14 Higashi-ikebukuro Toshima-Ku,Tokyo 170-0013 JAPAN

Manufacturer^b A&D Compnay,Limited Address 3-23-14 Higashi-ikebukuro Toshima-Ku,Tokyo 170-0013 JAPAN

Brand^c A&D Model^d TM-2440

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker^a A&D Compnay,Limited Address 3-23-14 Higashi-ikebukuro Toshima-Ku,Tokyo 170-0013 JAPAN

Manufacturer^b A&D Compnay,Limited Address 3-23-14 Higashi-ikebukuro Toshima-Ku,Tokyo 170-0013 JAPAN

Brand^c A&D Model^d TM-2441

Existing validated blood pressure measuring device.

which has previously passed the ISO81060-2 protocol, the results of which were published as follows:

Kario.K, Hoshide.S, et al. Validation of the TM-2441 ambulatory blood pressure measurement device according to the ISO 81060-2:2013 standard

Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1–18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^e <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^g <input type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

- Notes:
- a Provide the name and address of the actual maker of the device.
 - b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
 - c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
 - d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
 - e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
 - f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
 - g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

- 9)The equivalent device model name:TM-2440
- 10)Difference of case design and function. Both devices have the different casing.
- 11)The display of TM-2440 is only OLED (display of TM-2441 is OLED and LCD).
- 13)Difference of switch function, sensing acceleration, Bluetooth communication, etc.
- 16)Bluetooth communication can not be used.

SECTION C Please check that the following are included with the application

- A manual for the validated device
 - A manual for the device for which equivalence is being sought
 - Completed DET9 Form
 - An image of the device for which equivalence is being sought
 - An image of the screen layout of validated device*
 - An image of the screen layout of the device for which equivalence is being sought*
- * Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director Yasuhiko Shinozaki

Name Yasuhiko Shinozaki

Date 22 July 2019

Signature of Witness Shinobu Ozaki

Name Shinobu Ozaki

Address 3-23-14 Higashi-ikebukuro Toshima-Ku,Tokyo 170-0013 JAPAN

Company Stamp/Seal



Comparison of the A&D TM2440 with the A&D TM2441

Devices – Item 9	A&D TM2440	A&D TM2441
Pictures		
Display Image		
Validation		ISO80601-2:2013
Category	Ambulatory Blood Pressure Monitor	Ambulatory Blood Pressure Monitor
Casing – Item 10	<p><i>Dimensions</i> Approx : 95 [D] × 66 [W] × 24.5 [H] mm</p> <p><i>Ports</i> Cuff port USB connector port</p> <p><i>Features</i> START/STOP switch for ABPM EVENT switch</p>	<p><i>Dimensions</i> Approx : 95 [D] × 66 [W] × 24.5 [H] mm</p> <p><i>Ports</i> Cuff port USB connector port</p> <p><i>Features</i> START/STOP switch for ABPM START/STOP switch for Self Measurement EVENT switch Exchange AUTO switch between ABPM and Self Measurement</p>

Display – Item 11	<i>Type</i> OLED display (dot matrix)	<i>Type</i> OLED display (dot matrix) + liquid crystal display
Carrying/Mounting Facilities – Item 12	N/A	N/A
Software other than Algorithm – Item 13	Auto blood pressure measurement USB communication	Auto blood pressure measurement Self blood pressure measurement Changing between ABPM and Self Acceleration sensing and memory Temperature and Barometric pressure sensing and memory Bluetooth communication USB communication
Memory Capacity Item 14	<i>Number of stored measurements</i> 600 blood pressure measurements	<i>Number of stored measurements</i> 600 blood pressure measurements
Printing Facilities Item 15	N/A	N/A
Communication Facilities – Item 16	USB communication	Bluetooth® Ver.4.1 Low Energy USB communication
Power Supply Item 17	2×1.5V batteries(LR6 or AA) Alkaline battery or Nickel-hydrogen battery 1900mAh or more	2×1.5V batteries(LR6 or AA) Alkaline battery or Nickel-hydrogen battery 1900mAh or more
Other differences	<i>Other Details on Equivalent device that are different to Validated device</i> <i>Sensors</i> N/A	<i>Other Details on Validated device that are different to Equivalent device</i> <i>Sensors</i> N/A
Same Criteria	Measurement <i>Sensors</i> Semiconductor sensor <i>Accuracy</i> Pressure: ±3 mmHg Pulse: ±5 % <i>Method</i> Oscillometric measurement	Measurement <i>Sensors</i> Semiconductor sensor <i>Accuracy</i> Pressure: ±3 mmHg Pulse: ±5 % <i>Method</i> Oscillometric measurement

	<p><i>Ranges</i> Pressure: 0 - 299 mmHg Systolic pressure: 60 - 280 mmHg Diastolic pressure: 30 - 160 mmHg Pulse: 30 - 200 beats/minute</p> <p><i>Inflation</i> Rolling pump</p> <p><i>Deflation</i> Electromagnetic constant exhaust valve</p> <p><i>Cuffs (Please state sizes and materials used)</i> Small cuff: 15cm-22cm Nylon Adult cuff: 20cm-31cm Nylon Large cuff: 36cm-50cm Nylon Extra large cuff: 36cm-50cm Nylon</p> <p><i>Measurement Records</i> Date&Time, SYS,DIA,PUL</p> <p><i>Measurements other than Blood Pressure</i> Pulse rate(PUL)</p> <p>Buttons/Switches</p> <p><i>Power</i> N/A</p> <p><i>Analysis</i> N/A</p> <p><i>Function</i> EVENT switch</p> <p><i>Event Marking</i> N/A</p> <p><i>Communication</i> N/A</p>	<p><i>Ranges</i> Pressure: 0 - 299 mmHg Systolic pressure: 60 - 280 mmHg Diastolic pressure: 30 - 160 mmHg Pulse: 30 - 200 beats/minute</p> <p><i>Inflation</i> Rolling pump</p> <p><i>Deflation</i> Electromagnetic constant exhaust valve</p> <p><i>Cuffs(Please state sizes and materials used)</i> Small cuff: 15cm-22cm Nylon Adult cuff: 20cm-31cm Nylon Large cuff: 36cm-50cm Nylon Extra large cuff: 36cm-50cm Nylon</p> <p><i>Measurement Records</i> Date&Time, SYS,DIA,PUL</p> <p><i>Measurements other than Blood Pressure</i> Pulse rate(PUL)</p> <p>Buttons/Switches</p> <p><i>Power</i> N/A</p> <p><i>Analysis</i> N/A</p> <p><i>Function</i> EVENT switch</p> <p><i>Event Marking</i> N/A</p> <p><i>Communication</i> N/A</p>
--	--	---

	<p>Display/Symbols/Indicators</p> <p><i>Preparation</i></p> <p>Positioning indicator Zero is blinking</p> <p><i>Date and Time</i></p> <p>Year, Month, Day, Hour, Minute</p> <p><i>Power</i></p> <p>Battery detection symbol</p> <p><i>Post Measurement</i></p> <p>Systolic blood pressure Diastolic blood pressure Pulse rate Date and Time</p> <p><i>Measurement Records</i></p> <p>Systolic blood pressure Diastolic blood pressure Pulse rate Date and Time Memory number</p> <p><i>Not described</i></p> <p>N/A</p> <p>Algorithms</p> <p><i>Averages and Differences</i></p> <p>N/A</p> <p><i>Diagnostic</i></p> <p>N/A</p>	<p>Display/Symbols/Indicators</p> <p><i>Preparation</i></p> <p>Positioning indicator Zero is blinking</p> <p><i>Date and Time</i></p> <p>Year, Month, Day, Hour, Minute</p> <p><i>Power</i></p> <p>Battery detection symbol</p> <p><i>Post Measurement</i></p> <p>Systolic blood pressure Diastolic blood pressure Pulse rate Date and Time</p> <p><i>Measurement Records</i></p> <p>Systolic blood pressure Diastolic blood pressure Pulse rate Date and Time Memory number</p> <p><i>Not described</i></p> <p>N/A</p> <p>Algorithms</p> <p><i>Averages and Differences</i></p> <p>N/A</p> <p><i>Diagnostic</i></p> <p>N/A</p>
<p>Comparable Criteria</p>	<p>Buttons/Switches</p> <p><i>Measurement Records</i></p> <p>START/STOP switch for ABPM</p>	<p>Buttons/Switches</p> <p><i>Measurement Records</i></p> <p>START/STOP switch for ABPM START/STOP switch for Self Measurement</p>

	<p>Display/Symbols/Indicators <i>Measurement Procedure</i> Pressure value</p> <p><i>Features</i> N/A</p> <p><i>Function</i> ABPM setting</p> <p><i>Communication</i> USB communication symbol</p> <p>Algorithms <i>Functions</i> IHB detection</p> <p><i>Communication</i> USB communication</p>	<p>Display/Symbols/Indicators <i>Measurement Procedure</i> Pressure value Heart mark (on LCD)</p> <p><i>Features</i> Display the minutes until the measurement (on LCD)</p> <p><i>Function</i> ABPM setting Self Measurement setting</p> <p><i>Communication</i> USB communication symbol (on LCD) Bluetooth communication symbol (on LCD)</p> <p>Algorithms <i>Functions</i> IHB detection (display on LCD)</p> <p><i>Communication</i> Bluetooth® Ver.4.1 Low Energy USB communication</p>
--	--	--

Comments	
Recommendation	Recommended
Date	12th August 2019