

**DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2013**

A SIGNED COPY WILL BE POSTED ON THE [www.dableducational.org](http://www.dableducational.org) WEBSITE

**SECTION A - Please complete all items.**

I **Minoru Yoshimura,** a Director of **OMRON Healthcare Europe B.V.,**  
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

**Maker<sup>a</sup>** **OMRON Healthcare Co., Ltd.** **Address** **53 Kunotsubo, Terado-cho, Muko, Kyoto 617-0002, Japan**  
**Manufacturer<sup>b</sup>** **OMRON Healthcare Co., Ltd** **Address** **53 Kunotsubo, Terado-cho, Muko, Kyoto 617-0002, Japan**  
**Brand<sup>c</sup>** **OMRON** **Model<sup>d</sup>** **RS1 (HEM-6120-E)**

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

**Maker<sup>a</sup>** **OMRON Healthcare Co., Ltd.** **Address** **53 Kunotsubo, Terado-cho, Muko, Kyoto 617-0002, Japan**  
**Manufacturer<sup>b</sup>** **OMRON Healthcare Co., Ltd.** **Address** **53 Kunotsubo, Terado-cho, Muko, Kyoto 617-0002, Japan**  
**Brand<sup>c</sup>** **OMRON** **Model<sup>d</sup>** **RS3 (HEM-6130-E)**

Existing validated blood pressure measuring device.

which has previously passed the **ESH-IP** protocol, the results of which were published as follows:

Takahashi H, Yokoi T, Yoshika M. Validation of the OMRON RS3 (HEM-6130-E) wrist blood pressure monitor, in oscillometry mode, for clinic use and self measurement in a general population, according to the European Society of Hypertension International Protocol revision 2010 [Internet]. Dublin: dablEducational Trust; 2013 Feb 01 [cited 2013 Feb 14]. 4 p. Available from: [http://www.dableducational.org/Publications/2013/ESH-IP 2010 Validation of Omron RS3 \(HEM-6130-E\).pdf](http://www.dableducational.org/Publications/2013/ESH-IP 2010 Validation of Omron RS3 (HEM-6130-E).pdf)

Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1–18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>e</sup> <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>g</sup> <input type="checkbox"/>

**An explanation of each item ticked “Yes” must be included in Section B or on a separate sheet.**

- Notes:
- a Provide the name and address of the actual maker of the device.
  - b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
  - c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
  - d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
  - e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
  - f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
  - g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

**SECTION B** An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

9. Model name RS1 (HEM-6120-E)

10. No Memory button and Date/Time setting button.

11. No Average value symbol, Date/Time display, Blood pressure level indicator, Movement error symbol and Irregular heartbeat symbol.

13. No function of Average value, Date/Time, Movement error and Irregular heartbeat detection.

14. Last measurement only.

**SECTION C** Please check that the following are included with the application

A manual for the validated device

A manual for the device for which equivalence is being sought

An image of the validated device

An image of the device for which equivalence is being sought

An image of the screen layout of validated device\*

An image of the screen layout of the device for which equivalence is being sought\*

\* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

**SECTION D** Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to [info@dableducational.org](mailto:info@dableducational.org).

Signature of Director 

Company Stamp/Seal

Name Minoru Yoshimura

Date 14 Feb 2013





Signature of Witness 

Name Tomohiro Kukita

Address Scorpis 33, 2132 LR Hoofddorp, The Netherlands

OMRON HEALTHCARE EUROPE BV  
Scorpis 33  
NL-2132 LR Hoofddorp  
P.O.BOX 2050 NL-2130 GL Hoofddorp  
TEL +31-23 5544700  
FAX +31-23 5544701

Comparison of the Omron RS1 (HEM-6120-E) with the Omron RS3 (HEM-6130-E)

Devices	Omron RS1 (HEM-6120-E)	Omron RS3 (HEM-6130-E)
Pictures		
Display		
Validation		ESH 2010
Device 1 Criteria		
Same Criteria	<p><b>Measurement</b></p> <p><i>Accuracy</i></p> <p>BP accuracy ± 3 mmHg 1, 5</p> <p>Pulse accuracy ± 5% 1, 5</p> <p><i>Method</i></p> <p>Oscillometric measurement method 1, 5</p> <p>Pulse 40 bpm to 180 bpm 1, 5, 8</p> <p>Manually initiated measurements 13</p> <p>Measurements are from single inflations 13</p> <p><i>Inflation</i></p> <p>Inflation 0 mmHg to 299 mmHg 1, 5, 7</p> <p>Automatic Inflation 7</p>	<p><b>Measurement</b></p> <p><i>Accuracy</i></p> <p>BP accuracy ± 3 mmHg 1, 5</p> <p>Pulse accuracy ± 5% 1, 5</p> <p><i>Method</i></p> <p>Oscillometric measurement method 1, 5</p> <p>Pulse 40 bpm to 180 bpm 1, 5, 8</p> <p>Manually initiated measurements 13</p> <p>Measurements are from single inflations 13</p> <p><i>Inflation</i></p> <p>Inflation 0 mmHg to 299 mmHg 1, 5, 7</p> <p>Automatic Inflation 7</p>

Devices	Omron RS1 (HEM-6120-E)	Omron RS3 (HEM-6130-E)
<b>Same Criteria</b>	<p><b>Measurement</b></p> <p><i>Deflation</i></p> <p>Automatic Deflation 8</p> <p><i>Cuffs</i></p> <p>Wrist circ. ~ 13.5 cm to ~ 21.5 cm 6</p> <p><b>Buttons/Switches</b></p> <p><i>Power</i></p> <p>On/Off with Start/Stop (Start/Stop Label) 10</p> <p><b>Display/Symbols/Indicators</b></p> <p><i>Preparation</i></p> <p>Correct cuff wrapping indicator 11, 13, 18</p> <p><i>Measurement Procedure</i></p> <p>Deflation symbol 11</p> <p>During Measurement: BP Level &amp; Heartbeat 11</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>Measurement error <math>E_1, E_3, E_4, E_5, E_r</math> 11</p> <p><i>Measurement Records</i></p> <p>Memory icon 11</p> <p><i>Power</i></p> <p>Low battery 11, 17</p> <p><b>Algorithms</b></p> <p><i>Parameter Settings</i></p> <p>Correct cuff wrapping detection 13</p> <p><b>Case</b></p> <p><i>Display</i></p> <p>Single screen display 10</p> <p>Segment LCD 10</p> <p><i>Power</i></p> <p>2 “AAA” batteries ~ 300 measurements 17</p> <p>Automatic switch-off when not used for 2 min 17</p>	<p><b>Measurement</b></p> <p><i>Deflation</i></p> <p>Automatic Deflation 8</p> <p><i>Cuffs</i></p> <p>Wrist circ. ~ 13.5 cm to ~ 21.5 cm 6</p> <p><b>Buttons/Switches</b></p> <p><i>Power</i></p> <p>On/Off with Start/Stop (Start/Stop Label) 10</p> <p><b>Display/Symbols/Indicators</b></p> <p><i>Preparation</i></p> <p>Correct cuff wrapping indicator 11, 13, 18</p> <p><i>Measurement Procedure</i></p> <p>Deflation symbol 11</p> <p>During Measurement: BP Level &amp; Heartbeat 11</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>Measurement error <math>E_1, E_3, E_4, E_5, E_r</math> 11</p> <p><i>Measurement Records</i></p> <p>Memory icon 11</p> <p><i>Power</i></p> <p>Low battery 11, 17</p> <p><b>Algorithms</b></p> <p><i>Parameter Settings</i></p> <p>Correct cuff wrapping detection 13</p> <p><b>Case</b></p> <p><i>Display</i></p> <p>Single screen display 10</p> <p>Segment LCD 10</p> <p><i>Power</i></p> <p>2 “AAA” batteries ~ 300 measurements 17</p> <p>Automatic switch-off when not used for 2 min 17</p>
<b>Comparable Criteria</b>	<p><b>Measurement</b></p> <p><i>Measurement Records</i></p> <p>Memory: 1 measurement 14</p>	<p><b>Measurement</b></p> <p><i>Measurement Records</i></p> <p>Memory: 60 measurements 14</p>

Devices	Omron RS1 (HEM-6120-E)	Omron RS3 (HEM-6130-E)
Device 2 Criteria		<p><b>Buttons/Switches</b></p> <p><i>Measurement Records</i></p> <p>Memory 10</p> <p><i>Settings</i></p> <p>Set 10</p> <p><b>Display/Symbols/Indicators</b></p> <p><i>Post Measurement</i></p> <p>Hypertension (Indicator strip) 11, 13</p> <p>BP classification (Thresholds exceeded) 10, 11, 13</p> <p>Average 11, 13, 14</p> <p>Body movement error 3, 11, 13, 18</p> <p>Irregular heartbeat 11, 13, 18</p> <p><i>Measurement Records</i></p> <p>Memory recall number (Replaces pulse rate momentarily) 11</p> <p><i>Date and Time</i></p> <p>Date and Time 11</p> <p>Date and Time (During memory recall) 11</p> <p><b>Algorithms</b></p> <p><i>Averages and Differences</i></p> <p>Last 3 measurements (within 10 min of each other) mean 13</p> <p><i>Diagnostic</i></p> <p>135 / 85 mmHg thresholds 13</p> <p>Irregular heartbeat detection 13</p> <p>Body movement error detection 3, 13</p>

Comments	1	Note	These devices are clearly equivalent and from the same family. The RS1 is a basic version with none of the extra features of the RS3 apart from the cuff wrapping indicator.
Recommendation	Equivalence is Recommended		
Date	15/02/2013		