

## DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2006

A SIGNED COPY WILL BE POSTED ON THE [www.dableducational.org](http://www.dableducational.org) WEBSITE

### SECTION A - Please complete all items online.

I Takefumi Nakanishi Director of Omron Healthcare Europe B.V.  
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Omron R6 (HEM-6000-E7)  
Blood pressure measuring device for which validation is claimed

blood pressure measuring device and the

Omron R7 (HEM-637-IT)  
Existing validated blood pressure measuring device

blood pressure measuring device, which has previously passed the International protocol, the results of which were published as follows

Topouchian JA, El Assaad MA, Orobinskaia LV, El Feghali RN, Asmar RG  
Authors(s)

Validation of two automatic devices for self-measurement of blood pressure according to the International Protocol of the European Society of Hypertension:

the Omron M6 (HEM-7001-E) and the Omron R7 (HEM 637-IT)

Title  
Blood Pressure Monitoring 2006; 11: 165-171  
Publication Year Volume Pages

The only differences between the devices involve the following components:

(When a component is not relevant, both Yes and No should be left blank. Please provide details on any differences below.)

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	5	Pressure Transducer	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	6	Cuff or Bladder	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	14	Memory Capacity/Number of stored measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	18	Other Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>

Brief explanation of differences and further relevant details:

5) The pressure sensor is replaced to a piezo electric sensor (NPS) from an electrostatic capacitive sensor (CPS), but the accuracy of blood pressure measurement is equivalent between NPS and CPS.

10) No "GRAPH" button, no two of "allow" buttons to select the setting and to view the graph display. No USB data communication port.

11) Segment LCD display instead of dot matrix LCD display, no graph display, no symbol of alarm function, includes the symbol of irregular heartbeat detection.

13) No graph display function, no alarm function, includes irregular heartbeat detection function, includes the memory average function (average of the latest 3 measurements).



16) No communication facilities.



**SECTION B** - Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original along with manuals for both devices to our address below.

Signature of Director	<u>T. Nakanishi</u>	Company Stamp/Seal
Name	<u>Takefumi Nakanishi</u>	OMRON HEALTHCARE EUROPE B.V.†
Date	<u>17 February 2010</u>	Kruisweg 577
Signature of Witness	<u>J. Meijer</u>	NL-2132 NA Hoofddorp
Name	<u>Janet Meijer</u>	P.O. Box 2150 NL- 2130 GL Hoofddorp
Address	<u>Omron Healthcare Europe B.V., Kruisweg 577, 2132NA Hoofddorp, The Netherlands</u>	Tel. +31 - 20 354 82 00
		Fax +31 - 20 354 82 01

Comparison of the Omron R6 (HEM-6000-E7) with the Omron R7 (HEM-637-IT)

Devices	Omron R6 (HEM-6000-E7)	Omron R7 (HEM-637-IT)
Pictures	 <p data-bbox="667 539 949 563">Image Omron R6 (HEM-6000-E)</p>	
Validation		ESH
Device 1 Criteria	<p data-bbox="387 635 663 659"><b>Display/Symbols/Indicators</b></p> <p data-bbox="387 663 618 687"><i>Measurement Procedure</i></p> <p data-bbox="416 692 645 716">Irregular heartbeat 11, 13</p> <p data-bbox="416 721 517 745">Average 11, 13</p> <p data-bbox="387 750 472 774"><i>Settings</i></p> <p data-bbox="416 778 546 802">Right wrist 11</p> <p data-bbox="387 807 499 831"><b>Algorithms</b></p> <p data-bbox="387 836 483 860"><i>Averages</i></p> <p data-bbox="416 865 745 888">Last 3 measurements mean 13</p> <p data-bbox="387 893 495 917"><i>Diagnostic</i></p> <p data-bbox="416 922 730 946">Atrial fibrillation detection 13</p>	
Same Criteria	<p data-bbox="387 997 528 1021"><b>Measurement</b></p> <p data-bbox="387 1026 483 1050"><i>Accuracy</i></p> <p data-bbox="416 1054 689 1078">BP accuracy ± 3 mmHg 1, 5</p> <p data-bbox="416 1083 656 1107">Pulse accuracy ± 5% 1, 5</p> <p data-bbox="387 1112 472 1136"><i>Method</i></p> <p data-bbox="416 1141 846 1165">Oscillometric measurement method 1, 5</p> <p data-bbox="416 1169 696 1193">Pulse 40 bpm -180 bpm 1, 5</p> <p data-bbox="416 1198 898 1222">Measurements are from single inflations 13</p> <p data-bbox="416 1227 819 1251">Manually initiated measurements 13, 14</p> <p data-bbox="387 1256 472 1279"><i>Inflation</i></p> <p data-bbox="416 1284 779 1308">Inflation 0 mmHg - 299 mmHg 1, 5, 7</p> <p data-bbox="416 1313 645 1337">Automatic Inflation 7</p>	<p data-bbox="1254 997 1395 1021"><b>Measurement</b></p> <p data-bbox="1254 1026 1350 1050"><i>Accuracy</i></p> <p data-bbox="1283 1054 1556 1078">BP accuracy ± 3 mmHg 1, 5</p> <p data-bbox="1283 1083 1523 1107">Pulse accuracy ± 5% 1, 5</p> <p data-bbox="1254 1112 1339 1136"><i>Method</i></p> <p data-bbox="1283 1141 1713 1165">Oscillometric measurement method 1, 5</p> <p data-bbox="1283 1169 1563 1193">Pulse 40 bpm -180 bpm 1, 5</p> <p data-bbox="1283 1198 1765 1222">Measurements are from single inflations 13</p> <p data-bbox="1283 1227 1686 1251">Manually initiated measurements 13, 14</p> <p data-bbox="1254 1256 1339 1279"><i>Inflation</i></p> <p data-bbox="1283 1284 1646 1308">Inflation 0 mmHg - 299 mmHg 1, 5, 7</p> <p data-bbox="1283 1313 1512 1337">Automatic Inflation 7</p>

Devices	Omron R6 (HEM-6000-E7)	Omron R7 (HEM-637-IT)
Same Criteria (Continued)	<b>Measurement</b>	<b>Measurement</b>
	<i>Deflation</i>	<i>Deflation</i>
	Automatic Deflation	Automatic Deflation 8
	<i>Cuffs</i>	<i>Cuffs</i>
	Wrist circ. 13.5-21.5 cm	Wrist circ. 13.5-21.5 cm 6
	<i>Sensors</i>	<i>Sensors</i>
	Wrist positioning sensor	Wrist positioning sensor 13, 18
	<i>Measurement Records</i>	<i>Measurement Records</i>
	Memory: 90 measurements	Memory: 90 measurements 14
	<b>Buttons/Switches</b>	<b>Buttons/Switches</b>
	<i>Measurement Records</i>	<i>Measurement Records</i>
	Memory	Memory 10
	<i>Settings</i>	<i>Settings</i>
	Set	Set 10
	<b>Display/Symbols/Indicators</b>	<b>Display/Symbols/Indicators</b>
	<i>Measurement Procedure</i>	<i>Measurement Procedure</i>
	Deflation symbol	Deflation symbol 11
	Heartbeat symbol during deflation	Heartbeat symbol during deflation 11
	Wrist position – adjust and OK	Wrist position – adjust and OK 11, 13, 18
	<i>Post Measurement</i>	<i>Post Measurement</i>
	SBP, DBP and Pulse	SBP, DBP and Pulse 11
	<i>Date and Time</i>	<i>Date and Time</i>
	Date and Time	Date and Time 11
<i>Power</i>	<i>Power</i>	
Low battery	Low battery 11, 17	
<b>Algorithms</b>	<b>Algorithms</b>	
<i>Parameter Settings</i>	<i>Parameter Settings</i>	
Right or left wrist	Right or left wrist 1	
<b>Case</b>	<b>Case</b>	
<i>Display</i>	<i>Display</i>	
Single screen display	Single screen display 10	
<i>Power</i>	<i>Power</i>	
2 “AAA” batteries ~ 300 measurements	2 “AAA” batteries ~ 300 measurements 17	
Automatic switch-off when not used for 2 min	Automatic switch-off when not used for 2 min 17	

Devices	Omron R6 (HEM-6000-E7)	Omron R7 (HEM-637-IT)
<b>Comparable Criteria</b>	<p><b>Measurement</b> <i>Sensors</i> Pressure sensor: Piezoelectric semiconductor 5</p> <p><b>Buttons/Switches</b> <i>Power</i> On/Off with Start/Stop (O/I Start Label) 10</p> <p><b>Display/Symbols/Indicators</b> <i>Post Measurement</i> Measurement error (E ▼ , E height symbol ▼ , E/E ▼ , E onn) 11</p> <p><i>Measurement Records</i> Memory icon 11</p> <p><b>Case</b> <i>Display</i> Segment LCD 10</p>	<p><b>Measurement</b> <i>Sensors</i> Pressure sensor: Electrostatic capacitive semiconductor 5</p> <p><b>Buttons/Switches</b> <i>Power</i> On/Off with Start/Stop (O/I Label) 10</p> <p><b>Display/Symbols/Indicators</b> <i>Post Measurement</i> Measurement error (E ▼ , E HEIGHT ▼ , E/E, E Onn) 11</p> <p><i>Measurement Records</i> Memory “M” symbol 11</p> <p><b>Case</b> <i>Display</i> Dot matrix LCD 10</p>
<b>Device 2 Criteria</b>		<p><b>Buttons/Switches</b> <i>Settings</i> Forward 10 Backward 10</p> <p><i>Analysis</i> Graph 10</p> <p><b>Display/Symbols/Indicators</b> <i>Measurement Procedure</i> Inflation symbol 11 Graph 11, 13</p> <p><i>Date and Time</i> Alarm reminder (2 alarms/day) 18</p> <p><i>Settings</i> Screen font size adjustment 11</p> <p><b>Case</b> <i>Ports</i> USB port, cable and PC software 16, 18</p>
<b>Web link</b>	(HEM-6000-E) <a href="http://www.omron-healthcare.com/sitepreview.php?SiteID=234">http://www.omron-healthcare.com/sitepreview.php?SiteID=234</a>	(HEM-637-E2) <a href="http://www.omron-healthcare.com/sitepreview.php?SiteID=235">http://www.omron-healthcare.com/sitepreview.php?SiteID=235</a>

<b>Comments</b>	<p>The Omron R6 (HEM-6000-E) was approved for equivalence to the Omron R7 on 17/12/2008. The Omron R6 (HEM-6000-E7) is identical to the HEM-600- E device except that the current pressure sensor (CPS), an electrostatic capacitive semiconductor type, is changed to a new pressure sensor (NPS), a piezoelectric semiconductor type. Omron supplied dabl® Educational with full details of tests carried out (in confidence), and a summary of these tests for circulation to the advisory board. Further clarification on a number of queries was requested and provided. Following a review of these documents, it was concluded that the change in sensor would not affect the accuracy of the device and equivalence was recommended and was approved by the board.</p> <p><u>Original Comments</u></p> <p>The R6 and R7 appear to differ only in the extra features provided.</p> <p>The forward and back control buttons in the R7 are mainly there to move between individual measurements in the graph though they are used when setting the date and time. As the R6 does not have a graph, these buttons are not needed; in the R6 and the Memory (Mem) and Set buttons have arrows beside them and they are used as “forward” and “back” buttons as necessary.</p> <p>The arm positioning and blood pressure measurement instructions are almost identical for both devices. They just differ in the display details. The segment display in the R6 also allows icons to be displayed more readily. The cuffs are the same.</p>
<b>Recommendation</b>	Equivalence is approved.
<b>Date</b>	09/04/2010