

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2006

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items online.

I, Albert Lee Director of Health & Life Co., Ltd.
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Braun Exact Fit BP4900
Blood pressure measuring device for which validation is claimed

blood pressure measuring device and the

HL868BA
Existing validated blood pressure measuring device

blood pressure measuring device, which has previously passed the ESH protocol, the results of which were published as follows

Han-E Chen, Yan Cui, Chang-Sheng Sheng, Li-Hua Li, Yan Li, Ji-Guang Wang

Author(s)

Validation of the Healthy & Life HL868BA blood pressure monitor for home blood pressure

monitoring according to the European Society of Hypertension International Protocol

Title

Wolters Kluwer Health/Lippincott Williams & Wilkins

Publication

2008, Vol 13 No 5, Page 305-308

Year Volume Pages

The only differences between the devices involve the following components:

(When a component is not relevant, both Yes and No should be left blank. Please provide details on any differences below.)

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	6	Cuff or Bladder	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	18	Other Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>

Brief explanation of differences and further relevant details:

- Item 6 BP4900 use universal cuff (9"-17"). HL868BA use normal cuff (9"-13") and large cuff (13"-17"). Base on the test result from our Internal Clinical Report in the attached file, we think and confirm that our universal cuff does not impact the device accuracy.
- Item 9 Model name and number are different than HL868BA.
- Item 10 The device external and dimension are different than HL868BA.
- Item 11 LCD size and symbols are different than HL868BA.
- Item 13 BP4900 can calculate the full day average of last 7 days. HL868BA can calculate the average of last 3 memories.
- Item 14 BP4900 memory capacity 48 single user. HL868BA 3 user 80 memories each.
- Item 17 BP4900 power uses four AA batteries only. HL868BA can use battery or adaptor.
- Item 18 HL868BA has other facilities such as Irregular Heartbeat detector, Self-management, Temperature detect, and PC-Link. BP4900 has not foregoing facilities. Those other facilities don't affect the measurement algorithm.

dabl® Educational Trust

SECTION B - Complete all items, have signatures and seal, online and print. Sign and seal it then send the original along with manuals for both devices to our address below.

Signature of Director

Albert Lee

Company Stamp/Seal

Name

ALBERT LEE

Date

January 28, 2010

Signature of Witness

Bill Huang 2010-1-28



Name

BILL HUANG

Address

9F., No. 186, Jian Yi Road, Chung Ho City 23553, Taipei, Taiwan

Comparison of the Braun BP4900 with the Health and Life HL868BA

Devices	Braun BP4900	Health and Life HL868BA																																																																
Pictures																																																																		
Validation		ESH																																																																
Device 1 Criteria		<table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 80%;">Up and down buttons</td> <td style="text-align: right;">10</td> </tr> <tr> <td>Self diagnosis (Set thresholds, BPs flash if HBP, ☺ if OK)</td> <td style="text-align: right;">10, 11, 13</td> </tr> <tr> <td>Multi-measurements (3 or 5 continuous measurements)</td> <td style="text-align: right;">10, 11, 13</td> </tr> <tr> <td>Irregular heartbeat detection & symbol</td> <td style="text-align: right;">11, 13</td> </tr> <tr> <td>3-measurement or 5-measurement median</td> <td style="text-align: right;">13</td> </tr> <tr> <td>Power: Optional AC adapter</td> <td style="text-align: right;">17</td> </tr> </table>	Up and down buttons	10	Self diagnosis (Set thresholds, BPs flash if HBP, ☺ if OK)	10, 11, 13	Multi-measurements (3 or 5 continuous measurements)	10, 11, 13	Irregular heartbeat detection & symbol	11, 13	3-measurement or 5-measurement median	13	Power: Optional AC adapter	17																																																				
Up and down buttons	10																																																																	
Self diagnosis (Set thresholds, BPs flash if HBP, ☺ if OK)	10, 11, 13																																																																	
Multi-measurements (3 or 5 continuous measurements)	10, 11, 13																																																																	
Irregular heartbeat detection & symbol	11, 13																																																																	
3-measurement or 5-measurement median	13																																																																	
Power: Optional AC adapter	17																																																																	
Same Criteria	<table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 80%;">Accuracy ± 3 mmHg</td> <td style="text-align: right;">1, 5</td> </tr> <tr> <td>Oscillometric measurement method</td> <td style="text-align: right;">1, 5</td> </tr> <tr> <td>BP 0 mmHg to 300 mmHg, Pulse 40-199 bpm</td> <td style="text-align: right;">1, 5, 7, 8</td> </tr> <tr> <td>Semiconductor pressure sensor*</td> <td style="text-align: right;">5</td> </tr> <tr> <td>Automatic Inflation and Deflation</td> <td style="text-align: right;">7, 8</td> </tr> <tr> <td>Deflation: Automatic exhaust valve</td> <td style="text-align: right;">8</td> </tr> <tr> <td>Mode button</td> <td style="text-align: right;">10</td> </tr> <tr> <td>Single screen LCD display</td> <td style="text-align: right;">10</td> </tr> <tr> <td>Start/Stop button</td> <td style="text-align: right;">10</td> </tr> <tr> <td>Memory button, memory symbol</td> <td style="text-align: right;">10, 11</td> </tr> <tr> <td>Date/Time Display</td> <td style="text-align: right;">11</td> </tr> <tr> <td>During Measurement: Inflation, Deflation & Heartbeat Symbols</td> <td style="text-align: right;">11</td> </tr> <tr> <td>Measurement error symbol</td> <td style="text-align: right;">11</td> </tr> <tr> <td>Low battery symbol</td> <td style="text-align: right;">11, 17</td> </tr> <tr> <td>Power: 4 “AA” batteries ~ 300 measurements</td> <td style="text-align: right;">17</td> </tr> <tr> <td>Power: Automatic switch-off when not used for 1 min</td> <td style="text-align: right;">17</td> </tr> </table>	Accuracy ± 3 mmHg	1, 5	Oscillometric measurement method	1, 5	BP 0 mmHg to 300 mmHg, Pulse 40-199 bpm	1, 5, 7, 8	Semiconductor pressure sensor*	5	Automatic Inflation and Deflation	7, 8	Deflation: Automatic exhaust valve	8	Mode button	10	Single screen LCD display	10	Start/Stop button	10	Memory button, memory symbol	10, 11	Date/Time Display	11	During Measurement: Inflation, Deflation & Heartbeat Symbols	11	Measurement error symbol	11	Low battery symbol	11, 17	Power: 4 “AA” batteries ~ 300 measurements	17	Power: Automatic switch-off when not used for 1 min	17	<table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 80%;">Accuracy ± 3 mmHg</td> <td style="text-align: right;">1, 5</td> </tr> <tr> <td>Oscillometric measurement method</td> <td style="text-align: right;">1, 5</td> </tr> <tr> <td>BP 0 mmHg to 300 mmHg, Pulse 40-199 bpm</td> <td style="text-align: right;">1, 5, 7, 8</td> </tr> <tr> <td>Semiconductor pressure sensor*</td> <td style="text-align: right;">5</td> </tr> <tr> <td>Automatic Inflation and Deflation</td> <td style="text-align: right;">7, 8</td> </tr> <tr> <td>Deflation: Automatic exhaust valve</td> <td style="text-align: right;">8</td> </tr> <tr> <td>Mode button</td> <td style="text-align: right;">10</td> </tr> <tr> <td>Single screen LCD display</td> <td style="text-align: right;">10</td> </tr> <tr> <td>Start/Stop button</td> <td style="text-align: right;">10</td> </tr> <tr> <td>Memory button, memory symbol</td> <td style="text-align: right;">10, 11</td> </tr> <tr> <td>Date/Time Display</td> <td style="text-align: right;">11</td> </tr> <tr> <td>During Measurement: Inflation, Deflation & Heartbeat Symbols</td> <td style="text-align: right;">11</td> </tr> <tr> <td>Measurement error symbol</td> <td style="text-align: right;">11</td> </tr> <tr> <td>Low battery symbol</td> <td style="text-align: right;">11, 17</td> </tr> <tr> <td>Power: 4 “AA” batteries ~ 300 measurements</td> <td style="text-align: right;">17</td> </tr> <tr> <td>Power: Automatic switch-off when not used for 1 min</td> <td style="text-align: right;">17</td> </tr> </table>	Accuracy ± 3 mmHg	1, 5	Oscillometric measurement method	1, 5	BP 0 mmHg to 300 mmHg, Pulse 40-199 bpm	1, 5, 7, 8	Semiconductor pressure sensor*	5	Automatic Inflation and Deflation	7, 8	Deflation: Automatic exhaust valve	8	Mode button	10	Single screen LCD display	10	Start/Stop button	10	Memory button, memory symbol	10, 11	Date/Time Display	11	During Measurement: Inflation, Deflation & Heartbeat Symbols	11	Measurement error symbol	11	Low battery symbol	11, 17	Power: 4 “AA” batteries ~ 300 measurements	17	Power: Automatic switch-off when not used for 1 min	17
Accuracy ± 3 mmHg	1, 5																																																																	
Oscillometric measurement method	1, 5																																																																	
BP 0 mmHg to 300 mmHg, Pulse 40-199 bpm	1, 5, 7, 8																																																																	
Semiconductor pressure sensor*	5																																																																	
Automatic Inflation and Deflation	7, 8																																																																	
Deflation: Automatic exhaust valve	8																																																																	
Mode button	10																																																																	
Single screen LCD display	10																																																																	
Start/Stop button	10																																																																	
Memory button, memory symbol	10, 11																																																																	
Date/Time Display	11																																																																	
During Measurement: Inflation, Deflation & Heartbeat Symbols	11																																																																	
Measurement error symbol	11																																																																	
Low battery symbol	11, 17																																																																	
Power: 4 “AA” batteries ~ 300 measurements	17																																																																	
Power: Automatic switch-off when not used for 1 min	17																																																																	
Accuracy ± 3 mmHg	1, 5																																																																	
Oscillometric measurement method	1, 5																																																																	
BP 0 mmHg to 300 mmHg, Pulse 40-199 bpm	1, 5, 7, 8																																																																	
Semiconductor pressure sensor*	5																																																																	
Automatic Inflation and Deflation	7, 8																																																																	
Deflation: Automatic exhaust valve	8																																																																	
Mode button	10																																																																	
Single screen LCD display	10																																																																	
Start/Stop button	10																																																																	
Memory button, memory symbol	10, 11																																																																	
Date/Time Display	11																																																																	
During Measurement: Inflation, Deflation & Heartbeat Symbols	11																																																																	
Measurement error symbol	11																																																																	
Low battery symbol	11, 17																																																																	
Power: 4 “AA” batteries ~ 300 measurements	17																																																																	
Power: Automatic switch-off when not used for 1 min	17																																																																	
Comparable Criteria	<table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 80%;">Cuff: (Arm circ. 23 to 43 cm)</td> <td style="text-align: right;">6</td> </tr> <tr> <td>Memory: 48 measurements</td> <td style="text-align: right;">11, 14</td> </tr> </table>	Cuff: (Arm circ. 23 to 43 cm)	6	Memory: 48 measurements	11, 14	<table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 80%;">Cuff: (Arm circ. 23 to 33 cm)</td> <td style="text-align: right;">6</td> </tr> <tr> <td>Memory: 80 measurements × 3 users (zones)</td> <td style="text-align: right;">11, 14</td> </tr> </table>	Cuff: (Arm circ. 23 to 33 cm)	6	Memory: 80 measurements × 3 users (zones)	11, 14																																																								
Cuff: (Arm circ. 23 to 43 cm)	6																																																																	
Memory: 48 measurements	11, 14																																																																	
Cuff: (Arm circ. 23 to 33 cm)	6																																																																	
Memory: 80 measurements × 3 users (zones)	11, 14																																																																	
Device 2 Criteria	<table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 80%;">Average Button</td> <td style="text-align: right;">10</td> </tr> <tr> <td>Set Button</td> <td style="text-align: right;">10</td> </tr> <tr> <td>7-day average</td> <td style="text-align: right;">13</td> </tr> </table>	Average Button	10	Set Button	10	7-day average	13																																																											
Average Button	10																																																																	
Set Button	10																																																																	
7-day average	13																																																																	

Web link		
----------	--	--

Comments	<p>* Not in the manual but supplied on data sheets</p> <p>Eight queries were sent to the initial application, which included extra information. These are shown below with each Query, Response and Comment numbered according to the communication number. In three cases (#4, #6 & #7), the explanation was accepted. Two of the responses (#1 & #2) conflicted with the manuals and they were queried further. For the remaining three (#3, #5 & #8), a new corrected application was required. A new application was returned and was fine for all but for some minor errors.</p> <p>The company “Kaz”, mentioned in some responses, is the OEM manufacturer of both devices.</p> <p>1 Query 1 The difference in the <i>Printing Facilities</i> option (#15) is ticked as Yes in the application form. There do not appear to be printing facilities on either device.</p> <p>Response 1 <i>Printing facilities option (#15) was ticked off as “yes” because the printer which makes the owner’s manual and other packaging items is different than the HL868BA.</i></p> <p>Query 2 Can you please point out where the printing facilities are described in the manuals?</p> <p>Comment 2 The <i>Printing Facilities</i> option (#15) is ticked as <i>No</i> in the new application form.</p> <p>2 Query 1 The difference in the <i>Communication Facilities</i> option (#16) is ticked as Yes in the application form. Furthermore, a USB cable to link the HL868BA to a PC is stated in the additional information provided. Yet there is no information available in the manual.</p> <p>Response 1 <i>The USB cable or PC link feature is only available on H&L model HL868BA not available for Kaz models. This is why we do not mention it in the owner’s manual.</i></p> <p>Query 2 Where is the USB port in the HL868BA? It is not mentioned in the manual and there is no information available in the Health and Life or other websites.</p> <p>Comment 2 The <i>Communication Facilities</i> option (#16) is ticked as <i>No</i> in the new application form. However, “PC-Link” is erroneously included in as a difference under the <i>Other Facilities</i> option (#18). However, it is not a measurement or validation matter.</p> <p>3 Query 1 The difference in the <i>Power Supply</i> option (#17) is ticked as No in the application form. Yet there is an optional power supply with the HL868BA that is not provided with the BP4900.</p> <p>Response 1 <i>The Kaz models do not have an optional power supply. The power source for HL868BA and Kaz models are the same; both use 4 AA 1.5V alkaline batteries.</i></p> <p>Comment 1 Explanation accepted. However, this should have been ticked as “Yes” with the explanation that an optional AC adaptor is available only with the HL868BA.</p> <p> This must be corrected in a fresh application form.</p> <p>Comment 2 The <i>Power Supply</i> option (#17) is ticked as <i>Yes</i> in the new application form and the explanation is provided.</p> <p>4 Query 1 The measurement range for blood pressure, for both devices is 0-300 mmHg in the manuals but 40-280 mmHg in the additional information provided.</p> <p>Response 1 <i>These devices are designed to measure blood pressure values ranging 0-300 mmHg (range of inflation), and 40-280 mmHg is range of</i></p>	
----------	---	--

		<i>measurement.</i>
	Comment 1	Explanation accepted
5	Query 1	The "Average display" claimed, in the additional information provided, as a feature of the HL868BA, does not appear to be in the manual. A facility for measuring 3 or 5 measurements and returning the "middle one" (presumably the median) is described in the manual.
	Response 1	<i>The "Average display" for Kaz models is different than H&L's model as indicated in the application (#13 & #14). Kaz's "Average display" feature specifications are explained in the owner's manuals.</i>
	Comment 1	<p>Items #13 and #14 are ticked as "Yes" but the application does not provide a paragraph for each "Yes" clearly stating what the differences are. Instead, a matrix is provided from which one has to try and work out what which differences match which "Yes". In this case, there is a row in the matrix labelled "Average display" which is marked as "Last 7 days data average" for the BP4900 and "Last 3 memories" for the HL868BA.</p> <p>Although "average" can mean any type of "typical" measurement, it is usually taken to mean the "mean value". The manual for the HL868BA describes a "multi-measurement option in which 3 or 5 measurements can be taken automatically in succession and "after all measurements have been taken, the middle reading of each Systolic, Diastolic and Pulse reading will be displayed on the screen.". These "middle readings" presumably refer to the medians. Though medians are, indeed, averages, "Average display – last 3 memories" does not accurately describe this feature.</p> <p>On the other hand, the "correct 7-Day Average result", available in the BP4900, presumably means the arithmetic mean of the "full day averages" (also presumably the arithmetic mean of the BPs during that day) taken during the preceding 7 days.</p> <p>The devices provide very different averages both in kind and scope. This must be described correctly in a fresh application.</p>
	Comment 2	The response in the new application form states "BP4900 can calculate the full day average of last 7 days. HL868BA can calculate the average of last 3 memories". The median facility is not added. However, it is not a measurement or validation matter.
6	Query 1	The 12hr or 24hr option for the BP4900 clock claimed, in the additional information provided, does not appear to be in the manual.
	Response 1	<i>12hr or 24hr option is not mentioned in the owner's manual as this is pre-set at the factory (H&L). The user does not have the option to set 12hr or 24 hr. This option is for internal use only and is set at the factory pending the region the device will be sold. So all units shipping to USA will be pre-set to 12hr and all units shipping to Europe will be set at 24hrs.</i>
	Comment 1	Explanation accepted
7	Query 1	The Deflation time, Deflation rate of air leakage, and Specification of pressure do not appear to be in the manuals. From where does this information come?
	Response 1	<i>Our products were designed per EN1060-1 & EN1060-3 protocols, which are engineering specifications and aren't printed in the user's manual.</i>
	Comment 1	Explanation accepted
8	Query 1	The cuff size, in the HL868BA manual, is suitable for arm circumferences in the range 23-33 cm/9-13 inch, and a warning to this effect is included. However, the paper validating the HL868BA and the manual for the BP4600 state that a larger cuff, for arm circumferences 33-43 cm, is also available. The manual for the BP4900 and BP5900 states that a single 23-43 cm/9-17 inch cuff is provided. In the additional information provided, this cuff is declared for both devices. The HL868BA was not validated with this new cuff covering a wider arm

	<p><i>circumference range. The difference in the Cuff or Bladder option (#6) is ticked as No in the application form.</i></p> <p>Response 1 <i>I confirm the BP4900 and BP5900 are equipped with a single 9-17 inch cuff. Aside from the arm circumference, the cuff is equal to two approved cuffs in the ESH journal. Cuff/bladder option may have been ticked off incorrectly.</i></p> <p>Comment 1 The application must be corrected. The “No” in the application form is incorrect. This is an important difference. The large cuff was only used in one patient in the study. Furthermore, a single cuff covering a wide range of arm circumferences is not necessarily the same as separate cuffs each covering narrower ranges.</p> <p>Comment 2 A separate internal validation was carried out by Health and Life to validate the HL868BA with the universal cuff.</p>
Recommendation	Equivalence is approved
Date	25/06/2010