

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2006

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items online.

I Liu Yi Director of Andon Health Co., Ltd
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Andon KD-931
Blood pressure measuring device for which validation is claimed

blood pressure measuring device and the

Andon KD-5915
Existing validated blood pressure measuring device

blood pressure measuring device, which has previously passed the ESH protocol, the results of which were published as follows

Qi-Fang Huang, Jie Wang, Chang-Sheng Sheng, Na-Na Zhang, Yan Li and Ji-Guang Wang

Authors(s)
Validation of the ANDON KD-5915 blood pressure monitor for home blood pressure monitoring according to the European Society of Hypertension International Protocol


Title
Blood Pressure Monitoring 2010;volume15;pp232-234
Publication Year Volume Pages

The only differences between the devices involve the following components:

(When a component is not relevant, both Yes and No should be left blank. Please provide details on any differences below.)

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	6	Cuff or Bladder	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	14	Memory Capacity/Number of stored measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	16	Communication Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	17	Power Supply	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	18	Other Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>

Brief explanation of differences and further relevant details:

- (10) Just has 1 button <STOP> with LED lighting; an iPhone connector; a USB port for adapter power supply.
- (11) As a display device, the iPhone instead the traditional LCD.
- (13) No voice function.
- (14) Stores 1000*4 readings instead of 60 readings
- (16) Communication with iPhone
- (17) DC:5V  2.1A, batteries: 3.7V Li-ion 400mAh

dabl® Educational Trust


SECTION B - Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original along with manuals for both devices to our address below.

Signature of Director Liu Yi

Company Stamp/Seal

Name Liu yi

Date 6 September 2010





Signature of Witness Liu zhijun 

Name Liu zhijun

Address Andon Health Co.,Ltd . No.3 Jin Ping Street,Ya An Road,Nankai District,Tianjin 300790,China



Comparison of the Andon KD-931 with the Andon KD-5915

Devices	Andon KD-931	Andon KD-5915
Pictures		
Display		
Validation		ESH
Device 1 Criteria	<p>Buttons/Switches</p> <p><i>Power</i></p> <p>Stop button 10</p> <p>Display/Symbols/Indicators</p> <p><i>Preparation</i></p> <p>Error on iPhone (ERROR: Communication Error) 11, 16</p> <p>Case</p> <p><i>Ports</i></p> <p>iPhone connection port 15</p> <p><i>Power</i></p> <p>Charging/Charged LED 17</p>	
Same Criteria	<p>Measurement</p> <p><i>Accuracy</i></p> <p>BP accuracy ± 3 mmHg 1, 5</p>	<p>Measurement</p> <p><i>Accuracy</i></p> <p>BP accuracy ± 3 mmHg 1, 5</p>

Devices	Andon KD-931	Andon KD-5915
Same Criteria (continued)	<p>Measurement (continued)</p> <p><i>Method</i></p> <p>Oscillometric measurement method 1, 5</p> <p>Pulse 40 bpm -180 bpm 1, 5, 8</p> <p>Manually initiated measurements 13</p> <p>Measurements are from single inflations 13</p> <p><i>Inflation</i></p> <p>Automatic Inflation 7</p> <p>Inflation 0 mmHg - 295 mmHg 1, 5, 7</p> <p>Zero pressure check before inflation ^{Query 1} 7</p> <p><i>Deflation</i></p> <p>Automatic Deflation 8</p> <p><i>Cuffs</i></p> <p>Extra Large (Arm circ. 42-48 cm) (Optional) 6</p> <p>Large (Arm circ. 30-42 cm) (Optional) 6</p> <p>Medium (Arm circ. 22 to 30 cm) 6</p> <p>Display/Symbols/Indicators</p> <p><i>Measurement Procedure</i></p> <p>Beeps before measurement 18</p> <p>Case</p> <p><i>Power</i></p> <p>Automatic switch-off when not used for 1 min 17</p>	<p>Measurement (continued)</p> <p><i>Method</i></p> <p>Oscillometric measurement method 1, 5</p> <p>Pulse 40 bpm -180 bpm 1, 5, 8</p> <p>Manually initiated measurements 13</p> <p>Measurements are from single inflations 13</p> <p><i>Inflation</i></p> <p>Automatic Inflation 7</p> <p>Inflation 0 mmHg - 295 mmHg 1, 5, 7</p> <p>Zero pressure check before inflation 7</p> <p><i>Deflation</i></p> <p>Automatic Deflation 8</p> <p><i>Cuffs</i></p> <p>Extra Large (Arm circ. 42-48 cm) (Optional) 6</p> <p>Large (Arm circ. 30-42 cm) (Optional) 6</p> <p>Medium (Arm circ. 22 to 30 cm) 6</p> <p>Display/Symbols/Indicators</p> <p><i>Measurement Procedure</i></p> <p>Beeps before measurement 18</p> <p>Case</p> <p><i>Power</i></p> <p>Automatic switch-off when not used for 1 min 17</p>
Comparable Criteria	<p>Measurement</p> <p><i>Sensors</i></p> <p>Pressure sensor: KD-2107-006GA ^{Query 2} 5</p> <p>Case</p> <p><i>Power</i></p> <p>3.7V Li-ion 400 mAh battery (rechargeable) 17</p> <p>AC adapter 17</p>	<p>Measurement</p> <p><i>Sensors</i></p> <p>Pressure sensor: KD-2107-006G or KD-2107-006GR 5</p> <p>Case</p> <p><i>Power</i></p> <p>4 "AA" batteries 17</p> <p>AC adapter (Optional) 17</p>
iPhone-Device Same Criteria	<p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>Hypertension (Indicator strip) 11, 13</p> <p>BP classification (WHO) 10, 11, 13</p> <p>Irregular heartbeat 11, 13, 18</p>	<p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>Hypertension (Indicator strip) 11, 13</p> <p>BP classification (WHO) 10, 11, 13</p> <p>Irregular heartbeat 11, 13, 18</p>

Devices	Andon KD-931	Andon KD-5915
iPhone-Device Same Criteria (continued)	<p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>Hypertension (Indicator strip) 11, 13</p> <p>BP classification (WHO) 10, 11, 13</p> <p>Irregular heartbeat 11, 13, 18</p> <p><i>Measurement Records</i></p> <p>Delete memory 11</p> <p><i>Date and Time</i></p> <p>Date and Time 11</p> <p>Date and Time (During memory recall) 11</p> <p>Algorithms</p> <p><i>Diagnostic</i></p> <p>WHO Guidelines 13</p> <p>Irregular heartbeat detection 13</p>	<p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>Hypertension (Indicator strip) 11, 13</p> <p>BP classification (WHO) 10, 11, 13</p> <p>Irregular heartbeat 11, 13, 18</p> <p><i>Measurement Records</i></p> <p>Delete memory 11</p> <p><i>Date and Time</i></p> <p>Date and Time 11</p> <p>Date and Time (During memory recall) 11</p> <p>Algorithms</p> <p><i>Diagnostic</i></p> <p>WHO Guidelines 13</p> <p>Irregular heartbeat detection 13</p>
iPhone-Device Comparable Criteria	<p>Measurement (continued)</p> <p><i>Measurement Records</i></p> <p>Memory: 1000 measurements × 4 zones 14</p> <p>Buttons/Switches</p> <p><i>Power</i></p> <p>On/Off and Start 10</p> <p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p>Measurement error 11</p> <p>ERROR: Pressure system is unstable before measurement</p> <p>ERROR: Fail to detect systolic pressure</p> <p>ERROR: Fail to detect diastolic pressure</p> <p>ERROR: Pneumatic system blocked or cuff is too tight during inflation</p> <p>ERROR: Pneumatic system leakage or cuff is too loose during inflation</p> <p>ERROR: Cuff pressure above 300mmHg</p> <p>ERROR: More than 160 seconds with cuff pressure above 15 mmHg</p> <p>ERROR: EEPROM accessing error</p> <p>ERROR: Device parameter checking error</p> <p>ERROR: Span Error^{Query 3}</p> <p><i>Measurement Records</i></p> <p>“Database” option on iPhone 11</p>	<p>Measurement (continued)</p> <p><i>Measurement Records</i></p> <p>Memory: 60 measurements 14</p> <p>Buttons/Switches</p> <p><i>Power</i></p> <p>On/Off with Start/Stop (Start Label) 10</p> <p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p>Measurement error 11</p> <p>Er 0</p> <p>Er 1</p> <p>Er 2</p> <p>Er 3</p> <p>Er 4</p> <p>Er 5</p> <p>Er 6 (3 minutes instead of 160 seconds)</p> <p>Er 7</p> <p>Er 8</p> <p>Er A^{Query 3}</p> <p><i>Measurement Records</i></p> <p>Memory “M” symbol 11</p> <p>Memory recall number 11</p>

Devices	Andon KD-931	Andon KD-5915
<p>iPhone-Device Comparable Criteria (continued)</p>	<p>Measurement (continued) <i>Measurement Records</i> Memory: 1000 measurements × 4 zones 14</p> <p>Buttons/Switches <i>Power</i> On/Off and Start 10</p> <p>Display/Symbols/Indicators <i>Post Measurement</i> Measurement error 11 ERROR: Pressure system is unstable before measurement ERROR: Fail to detect systolic pressure ERROR: Fail to detect diastolic pressure ERROR: Pneumatic system blocked or cuff is too tight during inflation ERROR: Pneumatic system leakage or cuff is too loose during inflation ERROR: Cuff pressure above 300mmHg ERROR: More than 160 seconds with cuff pressure above 15 mmHg ERROR: EEPROM accessing error ERROR: Device parameter checking error ERROR: Span Error^{Query 3}</p> <p><i>Measurement Records</i> “Database” option on iPhone 11</p> <p><i>Power</i> Low battery with remaining capacity detail 11, 17 Measurement error (ERROR: Low battery) 11</p> <p>Case <i>Display</i> iPhone screen 10</p>	<p>Measurement (continued) <i>Measurement Records</i> Memory: 60 measurements 14</p> <p>Buttons/Switches <i>Power</i> On/Off with Start/Stop (Start Label) 10</p> <p>Display/Symbols/Indicators <i>Post Measurement</i> Measurement error 11 Er 0 Er 1 Er 2 Er 3 Er 4 Er 5 Er 6 (3 minutes instead of 160 seconds) Er 7 Er 8 Er A^{Query 3}</p> <p><i>Measurement Records</i> Memory “M” symbol 11 Memory recall number 11</p> <p><i>Power</i> Low battery symbol 11, 17</p> <p>Case <i>Display</i> Single screen display 10 Segment LCD 10</p>
<p>Device 2 Criteria</p>		<p>Buttons/Switches <i>Measurement Records</i> Memory 10</p> <p>Display/Symbols/Indicators <i>Measurement Procedure</i> Optional voiced assistance 18 <i>Post Measurement</i> Optional voiced results 18 <i>Measurement Records</i> Optional voiced records 18</p>

Devices	Andon KD-931	Andon KD-5915
Device 2 Criteria		<p>Display/Symbols/Indicators (continued)</p> <p><i>Settings</i></p> <p>Current unit (kPa / mmHg) marker ^{Query 4} 11</p> <p>Case</p> <p><i>Power</i></p> <p>Rechargeable batteries not permitted 17</p>
Web link		http://andon.manufacturer.globalsources.com/si/6008800800032/pd/tl/Wrist-blood/1032471456/Blood-Pressure-Monitor.htm

Comments	<p>There were four queries</p> <p>Query 1 On the display screen of the KD-5915, there is a symbol indicating a zero pressure check. Does such a check happen with the KD-931?</p> <p>Response 1 <i>“Zero pressure check” happens with the KD-931, but there is not a symbol indicating.</i></p> <p>Query 2 The pressure sensor on the KD-931 is the KD-2107-006GA but on the KD-5915 it is the KD-2107-006G or KD-2107-006GR. What is the difference between the KD-2107-006GA and the other two sensors?</p> <p>Response 2 <i>KD-2107-006GA and KD-2107-006G: The pins of KD-2107-006GA are silvered, but KD-2107-006G’s pins are not.</i> <i>KD-2107-006GA and KD-2107-006GR: The way of installation is different because of the pins’ position.</i></p> <p>Query 3 The KD-5915 has an error ER A “Pressure sensor parameter error”. There does not appear to be a corresponding error on the KD-931. Why not?</p> <p>The KD-931 has a “Span Error” that did not exist on the KD-5915. What causes this error?</p> <p>Response 3 <i>“Pressure sensor parameter error” is equal to “Span Error”.</i> <i>Span Value = Pressure sensor value on 300mmHg – Pressure sensor value on 0mmHg.</i> <i>If Span Value is too small, the precision will decline. If Pressure sensor parameter error, the span will error.</i></p> <p>Query 4 On the display screen of the KD-5915, there are units for mmHg and kPa that seem to indicate a conversion facility. No such ability is described. Can you please confirm this facility either way?</p> <p>Response 4 <i>When the “START” button is pressed, all display characters are shown for self-test. The kPa is only displayed at the moment. It’s a reserved function that the result is displayed for kPa. The function doesn’t open, so the Operation Guide of the KD-5915 doesn’t mention it.</i></p>
----------	---

	<p>Although this appears to be a very different device, the differences are purely on the presentation of the results. Andon provide hardware information on the pump, valve and sensor. The pumps and valves are the same for both devices and the differences in the sensors only concern the pin coatings and positions.</p> <p>The differences therefore concern the presentation and storage of the results. The KD-931 does not provide any display options on the device itself and it does not store the measurements. All that is provided is an interface for an Apple iPhone. An app for the iPhone is also provided. The iPhone must be sitting in the slot provided in the KD-931 with the app running for the device to work. Measurements are initiated from the iPhone and the result of the measurement is returned to it. An emergency stop button is provided on the KD-931. The facilities provided by the app and the iPhone screen allow more measurements to be stored, more analysis of the results, including plots, and clearer explanation of errors.</p> <p>The fundamental measurement functions of both devices appear to be identical. The hardware, cuff and the operating parameters are the same.</p>
Recommendation	Equivalence is recommended
Date	16/09/2010