

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2011

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I **Yasutomo Kimiura,**
Name of a Company Directora Director of **Japan Precision Instruments Inc.,**
Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Manufacturer **Nissei** Brand **Nissei** Model **DSK-1051**

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the

Manufacturer **Nissei** Brand **Nissei** Model **DSK-1031**

Existing validated blood pressure measuring device. If alternative model names are used, include all.

blood pressure measuring device, which has previously passed the [ESH-IP 2010](#) protocol, the results of which were published as follows:[de Greeff A, Shennan AH.](#)

Authors(s)

Validation of the Nissei DSK-1031 upper arm oscillometric blood pressure monitor intended for clinic use and self measurement in a general population, according to the European Society of Hypertension International Protocol revision 2010 [Online]

Title

Available from URL:

2011 Oct 28 [2012 Dec 18].

http://www.dableducational.org/pdfs/ESHIP_Online_Validations/11-003

Nissei DSK-1031.pdf

Publication

Year Volume Pages

The only differences between the devices involve the following components:

When a component is not relevant, both Yes and No should be left blank. It is necessary to provide details on each item ticked "Yes" in Section C or on a separate sheet.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	6	Cuff or Bladder	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	10	Casing	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	18	Other Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section C on the next page

SECTION B Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form together with manuals and images for both devices to info@dableducational.org.

Signature of Director 

Company Stamp/Seal

Name **Yasutomo Kimiura**

Japan Precision Instruments Inc.
2508-13, Nakago, Shibukawa,
Gunma 377-0293, Japan

Date **19.12.2012**

Signature of Witness _____

Name

Address


SECTION C An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

Please refer to a comparison table.

Comparison of the Nissei DSK-1051 with the Nissei DSK-1031

Devices	Nissei DSK-1051 (Device 1)	Nissei DSK-1031 (Device 2)
Pictures		
Validation	ESH 2010 Protocol	
Comparable Criteria		
Software Other Than Algorithm Display	For the following functions	
WHO Indication	○	Graphic LCD
Trend Display	○	X
Temperature H/L Symbol	○	X
Locking Key symbol		X
Reliability Symbol		LCD Symbol Indication
Memory function		
Morning/Evening Memory	○	X
Memory Banks & Readings	2 x 120	2 x 60
Memory Avarage Readings	Recent 3 times readings	All readings

Comparison of the Nissei DSK-1051 with the Nissei DSK-1031

Devices	Nissei DSK-1051	Nissei DSK-1031
Pictures		
Display		
Validation		ESH-IP 2010
Device 1 Criteria	<p>Display/Symbols/Indicators</p> <p><i>Measurement Procedure</i></p> <p>Pulse level indicator</p> <p><i>Post Measurement</i></p> <p>Morning/Evening mean</p> <p>Ambient Temperature outside accuracy limits warming</p> <p>Last 5 measurements trend</p> <p><i>Date and Time</i></p> <p>Alarm reminder (one per memory zone)</p> <p><i>Features</i></p> <p>Key Lock</p> <p>Algorithms</p> <p><i>Averages and Differences</i></p> <p>Last 3 measurements taken over 15 min memory zone mean</p> <p>Last 3 morning meas. taken over 15 min memory zone mean</p>	<p>11, 13</p> <p>11, 13</p> <p>11, 13</p> <p>11, 13</p> <p>18</p> <p>11, 18</p> <p>13</p> <p>13</p>

Devices	Nissei DSK-1051	Nissei DSK-1031
Device 1 Criteria (continued)	Algorithms (continued) <i>Averages and Differences (continued)</i> Last 3 evening meas. taken over 15 min memory zone mean	13
Same Criteria	Measurement <i>Accuracy</i> BP accuracy ± 3 mmHg Pulse accuracy ± 5% <i>Method</i> Oscillometric measurement method SBP 50 mmHg – 250 mmHg, DBP 40 mmHg – 180 mmHg Pulse 40 bpm – 160 bpm Manually initiated measurements Measurements are from single inflations Automatic repeat on movement error <i>Inflation</i> Inflation 3 mmHg – 300 mmHg Automatic Inflation Fuzzy Logic Zero pressure check before inflation <i>Deflation</i> Automatic Deflation Automatic safety release valve <i>Cuffs</i> Single Cuff: Arm circumference 22 cm – 42 cm (DSC-EP10) ^{Query 2} Buttons/Switches <i>Measurement Records</i> Memory/Zone × 2 <i>Settings</i> Set Display/Symbols/Indicators <i>Measurement Procedure</i> Beeps before measurement Inflation symbol Deflation symbol Audible pulse indicator during deflation During Measurement: BP Level & Heartbeat	Measurement <i>Accuracy</i> BP accuracy ± 3 mmHg Pulse accuracy ± 5% <i>Method</i> Oscillometric measurement method SBP 50 mmHg – 250 mmHg, DBP 40 mmHg – 180 mmHg Pulse 40 bpm – 160 bpm Manually initiated measurements Measurements are from single inflations Automatic repeat on movement error <i>Inflation</i> Inflation 3 mmHg – 300 mmHg Automatic Inflation Fuzzy Logic Zero pressure check before inflation <i>Deflation</i> Automatic Deflation Automatic safety release valve <i>Cuffs</i> Single Cuff: Arm circumference 22 cm – 42 cm (DSC-EP10) Buttons/Switches <i>Measurement Records</i> Memory/Zone × 2 <i>Settings</i> Set Display/Symbols/Indicators <i>Measurement Procedure</i> Beeps before measurement Inflation symbol Deflation symbol Audible pulse indicator during deflation During Measurement: BP Level & Heartbeat

Devices	Nissei DSK-1051	Nissei DSK-1031
Same Criteria (continued)	<p>Display/Symbols/Indicators (continued)</p> <p><i>Measurement Procedure (continued)</i></p> <p>Beeps after measurement 18</p> <p>Correct cuff wrapping indicator (OK symbol) 11, 13, 18</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>PP 11</p> <p>Measurement error Err-300, Err-1, Err-2, Err-3 11</p> <p>BP classification (WHO) 10, 11, 13</p> <p>Average A 11, 13, 14</p> <p>Body movement error (体動 symbol) 3, 11, 13, 18</p> <p>Irregular heartbeat 11, 13, 18</p> <p><i>Date and Time</i></p> <p>Date and Time 11</p> <p>Date and Time (During memory recall) 11</p> <p><i>Measurement Records</i></p> <p>Memory, number of stored measurements 11</p> <p>Memory recall number 11</p> <p>Memory zone 11</p> <p><i>Power</i></p> <p>Low battery 11, 17</p> <p>Algorithms</p> <p><i>Diagnostic</i></p> <p>WHO Guidelines 13</p> <p>Irregular heartbeat detection 13</p> <p>Body movement error detection 3, 13</p> <p>Measurement Reliability 13</p> <p><i>Parameter Settings</i></p> <p>Correct cuff wrapping detection 13</p> <p>Casing</p> <p><i>Display</i></p> <p>Single screen Segment LCD display 10</p> <p><i>Power</i></p> <p>4 "AA" batteries 17</p> <p>AC adapter 17</p> <p>Automatic switch-off when not used for 3 min 17</p>	<p>Display/Symbols/Indicators (continued)</p> <p><i>Measurement Procedure</i></p> <p>Beeps after measurement 18</p> <p>Correct cuff wrapping indicator (OK symbol) 11, 13, 18</p> <p><i>Post Measurement</i></p> <p>SBP, DBP and Pulse 11</p> <p>PP 11</p> <p>Measurement error Err-300, Err-1, Err-2, Err-3 11</p> <p>BP classification (WHO) 10, 11, 13</p> <p>Average A 11, 13, 14</p> <p>Body movement error (Leftmost symbol) 3, 11, 13, 18</p> <p>Irregular heartbeat 11, 13, 18</p> <p><i>Date and Time</i></p> <p>Date and Time 11</p> <p>Date and Time (During memory recall) 11</p> <p><i>Measurement Records</i></p> <p>Memory, number of stored measurements 11</p> <p>Memory recall number 11</p> <p>Memory zone 11</p> <p><i>Power</i></p> <p>Low battery 11, 17</p> <p>Algorithms</p> <p><i>Diagnostic</i></p> <p>WHO Guidelines 13</p> <p>Irregular heartbeat detection 13</p> <p>Body movement error detection 3, 13</p> <p>Measurement Reliability 13</p> <p><i>Parameter Settings</i></p> <p>Correct cuff wrapping detection 13</p> <p>Case</p> <p><i>Display</i></p> <p>Single screen Segment LCD display 10</p> <p><i>Power</i></p> <p>4 "AA" batteries 17</p> <p>AC adapter (Optional) 17</p> <p>Automatic switch-off when not used for 3 min 17</p>

Devices	Nissei DSK-1051	Nissei DSK-1031
Comparable Criteria	<p>Measurement <i>Measurement Records</i> Memory: 120 measurements × 2 zones</p> <p>Buttons/Switches <i>Power</i> On/Off with Start/Stop and Key Lock/Unlock (Start/Stop Label)</p> <p>Display/Symbols/Indicators <i>Post Measurement</i> Hypertension (Flashing SBP and DBP)</p> <p>Measurement Reliability (LED)</p>	<p>Measurement <i>Measurement Records</i> Memory: 60 measurements × 2 zones</p> <p>Buttons/Switches <i>Power</i> On/Off with Start/Stop (Start/Stop Label)</p> <p>Display/Symbols/Indicators <i>Post Measurement</i> Hypertension (Indicator squares)</p> <p>Measurement Reliability (Centre symbol)</p>
Device 2 Criteria		<p>Algorithms <i>Averages and Differences</i> Memory zone mean</p>

Comments	1	The DSK-1051 is designed for the Japanese market. Blood pressure detection is identical to that of the DSK-1031 but it provides more post-measurement analysis as described above. During measurement, the grid, on the right, shows the pulse level. After measurement, it shows a plot of the BP level along with that of the previous four measurements. Translation of Japanese symbols (top down): 測定時刻: Measurement time. 月: Month. 年: Year. 日: Day. 最高血圧: Systolic blood pressure. 拍/分: Beats/minute. 脈圧: Pulse pressure. 最低血圧: Diastolic blood pressure. 体動: Body motion.
	2	<p>Query Is the DSC-EP10 cuff, used with the DSK-1031, also used with the DSK-1051?</p> <p>Response Yes, the cuff using for DSK-1031 and DSK-1051 is same. (DSC-EP10 cuff).</p> <p>Comment The confirmation is accepted.</p>
	3	<p>Query The manuals do not provide views of the full screen layouts. Please provide complete display screens for both devices.</p> <p>Response Images were provided as shown.</p>
Recommendation	Equivalence is recommended	
Date	07/01/2013	